

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2795

Client ID : SSTDCCC040

Date Analyzed: 08/08/2025

Lab File ID: BP025326.D

Time Analyzed: 10:15

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	240709	7.802	976682	10.57	575605	14.42
UPPER LIMIT	481418	8.302	1953360	11.072	1151210	14.919
LOWER LIMIT	120355	7.302	488341	10.072	287803	13.919
EPA SAMPLE NO.						
01 COMP-1	369826	7.80	1487910	10.57	927347	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.