

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2795

Client ID: SSTDCCC040

Date Analyzed: 08/08/2025

Lab File ID: BP025326.D

Time Analyzed: 10:15

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1048510	17.225	989888	21.666	1196040	25.059
UPPER LIMIT	2097020	17.725	1979780	22.166	2392080	25.559
LOWER LIMIT	524255	16.725	494944	21.166	598020	24.559
EPA SAMPLE NO.						
COMP-1	1790890	17.23	1327270	21.66	1321270	25.05

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.