

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025325.D
Instrument ID: BNA_P

Contract: POWE02
SDG NO.: Q2795
DFTPP Injection Date: 08/08/2025
DFTPP Injection Time: 09:33

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.3) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	78.0
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.6 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025326.D	08/08/2025	10:15
COMP-1	Q2795-01	BP025333.D	08/08/2025	15:11