

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW072225

Review By	Semsettin Yesilyurt	Review On	7/24/2025 3:08:17 PM
Supervise By	Mahesh Dadoda	Supervise On	7/24/2025 3:08:52 PM
SubDirectory	VW072225	HP Acquire Method	MSVOA_W HP Processing Method 82w072225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134865 VP134866,VP134868,VP134869,VP134870,VP134873,VP134874
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133934 VP134875

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031890.D	22 Jul 2025 08:14		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VW031891.D	22 Jul 2025 09:05	Comp. #20 is on Linear Regression	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VW031892.D	22 Jul 2025 09:37	Comp. #13 is on Quadratic Regression	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VW031893.D	22 Jul 2025 10:17		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VW031894.D	22 Jul 2025 10:39		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VW031895.D	22 Jul 2025 11:18		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VW031896.D	22 Jul 2025 12:00		SY/MD	Ok,M
8	VIBLK	VIBLK	VW031897.D	22 Jul 2025 12:21		SY/MD	Ok
9	VSTDICV050	ICVVW072225	VW031898.D	22 Jul 2025 14:47		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW080725

Review By	Maresh Dadoda	Review On	8/8/2025 3:06:45 PM
Supervise By	Semsettin Yesilyurt	Supervise On	8/8/2025 3:19:01 PM
SubDirectory	VW080725	HP Acquire Method	MSVOA_W HP Processing Method 82w072225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135033
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135034,VP135035 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW032035.D	07 Aug 2025 08:05		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW032036.D	07 Aug 2025 10:08		SY/MD	Ok,M
3	VW0807SBL01	VW0807SBL01	VW032037.D	07 Aug 2025 11:10		SY/MD	Ok
4	VW0807SBS01	VW0807SBS01	VW032038.D	07 Aug 2025 12:20		SY/MD	Ok,M
5	VW0807SBSD01	VW0807SBSD01	VW032039.D	07 Aug 2025 12:43		SY/MD	Ok,M
6	Q2784-01	EO-01-080625	VW032040.D	07 Aug 2025 13:20	vial-B	SY/MD	Ok
7	Q2793-02	VNJ-231	VW032041.D	07 Aug 2025 13:42	vial-A Internal Standard fail	SY/MD	ReRun
8	Q2795-01	COMP-1	VW032042.D	07 Aug 2025 14:04	vial-A	SY/MD	Ok
9	Q2795-02	COMP-2	VW032043.D	07 Aug 2025 14:26	vial-A	SY/MD	Ok
10	Q2795-03	COMP-3	VW032044.D	07 Aug 2025 14:48	vial-A	SY/MD	Ok
11	Q2793-02RE	VNJ-231RE	VW032045.D	07 Aug 2025 15:10	vial-B Internal Standard fail	SY/MD	Confirms
12	Q2798-03	TP-6-VOC	VW032046.D	07 Aug 2025 15:31	vial-A	SY/MD	Ok
13	Q2800-01	VNJ-222	VW032047.D	07 Aug 2025 15:53	vial-A	SY/MD	Ok
14	Q2801-01	TR-04-080725	VW032048.D	07 Aug 2025 16:15	vial-A	SY/MD	Ok
15	VIBLK	VIBLK	VW032049.D	07 Aug 2025 16:37		SY/MD	Ok
16	VSTDCCC050	VSTDCCC050EC	VW032050.D	07 Aug 2025 16:58		SY/MD	Ok,M

M : Manual Integration