

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q2795
 Lab File ID: VW032036.D Date Analyzed: 08/07/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	218306	7.96	410147	8.84	372672	11.63
UPPER LIMIT	436612	8.459	820294	9.343	745344	12.129
LOWER LIMIT	109153	7.459	205074	8.343	186336	11.129
EPA SAMPLE NO.						
COMP-1	174583	7.97	355865	8.85	344678	11.63
COMP-2	180542	7.96	382094	8.85	370140	11.64
COMP-3	179400	7.97	378554	8.85	363624	11.63
VW0807SBL01	166168	7.95	366099	8.85	354996	11.63
VW0807SBS01	222702	7.97	401808	8.85	364622	11.63
VW0807SBSD01	226447	7.96	405839	8.85	359768	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.