



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Environmental Restoration, LLC

Project Name: Cooper Chemical - Long Valley NJ 2-COOP-ANS

Project # N/A

Order ID # Q2799

Test Name: Mercury, Metals ICP-TAL

A. Number of Samples and Date of Receipt:

1 Water samples were received on 07/14/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Mercury, Metals ICP-TAL and METALS-TAL. This data package contains results for Mercury, Metals ICP-TAL.

C. Analytical Techniques:

The analysis and digestion of Metals ICP-TAL was based on 200.7 and The analysis and digestion of Mercury was based on 245.1.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate (EFF-WASTE WATERMSD) analysis met criteria for all compounds except for Magnesium, Potassium and Zinc due to Chemical Interference during Digestion process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments: The temperature of the samples at the time of receipt was 24.2°C.

The Post Digest Spike (EFF-WASTE WATERA) analysis met criteria for all compounds except for Potassium due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____