

Hit Summary Sheet SW-846

SDG No.: Q2799 **Order ID:** Q2799
Client: Environmental Restoration, LLC **Project ID:** Cooper Chemical - Long Valley NJ 2-COOI

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : CC-071325-RW								
Q2799-01	CC-071325-RW	Water	Aluminum	127		15.4	50.0	ug/L
Q2799-01	CC-071325-RW	Water	Barium	66.8		9.75	50.0	ug/L
Q2799-01	CC-071325-RW	Water	Cadmium	11.1		0.31	3.00	ug/L
Q2799-01	CC-071325-RW	Water	Calcium	11700		86.4	1000	ug/L
Q2799-01	CC-071325-RW	Water	Chromium	2.91	J	0.53	5.00	ug/L
Q2799-01	CC-071325-RW	Water	Cobalt	4.46	J	1.16	15.0	ug/L
Q2799-01	CC-071325-RW	Water	Copper	80.2		1.89	10.0	ug/L
Q2799-01	CC-071325-RW	Water	Iron	225		18.0	50.0	ug/L
Q2799-01	CC-071325-RW	Water	Magnesium	357	J	104	1000	ug/L
Q2799-01	CC-071325-RW	Water	Manganese	34.5		2.46	10.0	ug/L
Q2799-01	CC-071325-RW	Water	Nickel	14.3	J	1.90	20.0	ug/L
Q2799-01	CC-071325-RW	Water	Potassium	43500		309	1000	ug/L
Q2799-01	CC-071325-RW	Water	Silver	11.2		0.81	5.00	ug/L
Q2799-01	CC-071325-RW	Water	Sodium	64800		367	1000	ug/L
Q2799-01	CC-071325-RW	Water	Thallium	37.8		2.51	20.0	ug/L
Q2799-01	CC-071325-RW	Water	Zinc	2460		2.00	20.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	Environmental Restoration, LLC	Date Collected:	07/14/25
Project:	Cooper Chemical - Long Valley NJ 2-COOP-ANS	Date Received:	07/14/25
Client Sample ID:	CC-071325-RW	SDG No.:	Q2799
Lab Sample ID:	Q2799-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	127		1	15.4	50.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-36-0	Antimony	2.72	U	1	2.72	25.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-38-2	Arsenic	1.85	U	1	1.85	10.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-39-3	Barium	66.8		1	9.75	50.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-41-7	Beryllium	0.27	U	1	0.27	3.00	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-43-9	Cadmium	11.1		1	0.31	3.00	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-70-2	Calcium	11700		1	86.4	1000	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-47-3	Chromium	2.91	J	1	0.53	5.00	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-48-4	Cobalt	4.46	J	1	1.16	15.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-50-8	Copper	80.2		1	1.89	10.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7439-89-6	Iron	225		1	18.0	50.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7439-92-1	Lead	1.21	U	1	1.21	6.00	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7439-95-4	Magnesium	357	JN	1	104	1000	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7439-96-5	Manganese	34.5		1	2.46	10.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7439-97-6	Mercury	0.027	U	1	0.027	0.20	ug/L	08/12/25 13:50	08/12/25 18:23	E245.1	
7440-02-0	Nickel	14.3	J	1	1.90	20.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-09-7	Potassium	43500	N	1	309	1000	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7782-49-2	Selenium	2.52	U	1	2.52	10.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-22-4	Silver	11.2		1	0.81	5.00	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-23-5	Sodium	64800		1	367	1000	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-28-0	Thallium	37.8		1	2.51	20.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-62-2	Vanadium	2.49	U	1	2.49	20.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	
7440-66-6	Zinc	2460	N	1	2.00	20.0	ug/L	08/11/25 12:55	08/12/25 13:28	EPA 200.7	

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	METALS-TAL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB11	Mercury	0.027	+/-0.2	U	0.20	CV	08/12/2025	17:40	LB136791

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB32	Mercury	0.027	+/-0.2	U	0.20	CV	08/12/2025	17:45	LB136791
CCB33	Mercury	0.027	+/-0.2	U	0.20	CV	08/12/2025	18:20	LB136791
CCB34	Mercury	0.027	+/-0.2	U	0.20	CV	08/12/2025	18:32	LB136791

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	08/12/2025	11:48	LB136789
	Antimony	6.76	+/-25	U	50.0	P	08/12/2025	11:48	LB136789
	Arsenic	5.12	+/-10	U	20.0	P	08/12/2025	11:48	LB136789
	Barium	14.6	+/-50	U	100	P	08/12/2025	11:48	LB136789
	Beryllium	0.56	+/-3	U	6.00	P	08/12/2025	11:48	LB136789
	Cadmium	0.50	+/-3	U	6.00	P	08/12/2025	11:48	LB136789
	Calcium	234	+/-1000	U	2000	P	08/12/2025	11:48	LB136789
	Chromium	2.12	+/-5	U	10.0	P	08/12/2025	11:48	LB136789
	Cobalt	2.26	+/-15	U	30.0	P	08/12/2025	11:48	LB136789
	Copper	4.60	+/-10	U	20.0	P	08/12/2025	11:48	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	11:48	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	11:48	LB136789
	Magnesium	244	+/-1000	U	2000	P	08/12/2025	11:48	LB136789
	Manganese	5.94	+/-10	U	20.0	P	08/12/2025	11:48	LB136789
	Nickel	3.06	+/-20	U	40.0	P	08/12/2025	11:48	LB136789
	Potassium	918	+/-1000	U	2000	P	08/12/2025	11:48	LB136789
	Selenium	9.64	+/-10	U	20.0	P	08/12/2025	11:48	LB136789
	Silver	1.62	+/-5	U	10.0	P	08/12/2025	11:48	LB136789
	Sodium	868	+/-1000	U	2000	P	08/12/2025	11:48	LB136789
	Thallium	4.38	+/-20	U	40.0	P	08/12/2025	11:48	LB136789
	Vanadium	6.26	+/-20	U	40.0	P	08/12/2025	11:48	LB136789
	Zinc	16.7	+/-20	U	40.0	P	08/12/2025	11:48	LB136789

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	13.2	+/-50	J	100	P	08/12/2025	12:54	LB136789
	Antimony	6.76	+/-25	U	50.0	P	08/12/2025	12:54	LB136789
	Arsenic	5.12	+/-10	U	20.0	P	08/12/2025	12:54	LB136789
	Barium	14.6	+/-50	U	100	P	08/12/2025	12:54	LB136789
	Beryllium	0.56	+/-3	U	6.00	P	08/12/2025	12:54	LB136789
	Cadmium	0.50	+/-3	U	6.00	P	08/12/2025	12:54	LB136789
	Calcium	234	+/-1000	U	2000	P	08/12/2025	12:54	LB136789
	Chromium	2.12	+/-5	U	10.0	P	08/12/2025	12:54	LB136789
	Cobalt	2.26	+/-15	U	30.0	P	08/12/2025	12:54	LB136789
	Copper	4.60	+/-10	U	20.0	P	08/12/2025	12:54	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	12:54	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	12:54	LB136789
	Magnesium	244	+/-1000	U	2000	P	08/12/2025	12:54	LB136789
	Manganese	5.94	+/-10	U	20.0	P	08/12/2025	12:54	LB136789
	Nickel	3.06	+/-20	U	40.0	P	08/12/2025	12:54	LB136789
	Potassium	918	+/-1000	U	2000	P	08/12/2025	12:54	LB136789
	Selenium	9.64	+/-10	U	20.0	P	08/12/2025	12:54	LB136789
	Silver	1.62	+/-5	U	10.0	P	08/12/2025	12:54	LB136789
	Sodium	868	+/-1000	U	2000	P	08/12/2025	12:54	LB136789
	Thallium	4.38	+/-20	U	40.0	P	08/12/2025	12:54	LB136789
	Vanadium	6.26	+/-20	U	40.0	P	08/12/2025	12:54	LB136789
	Zinc	16.7	+/-20	U	40.0	P	08/12/2025	12:54	LB136789
CCB02	Aluminum	11.3	+/-50	U	100	P	08/12/2025	14:31	LB136789
	Antimony	6.76	+/-25	U	50.0	P	08/12/2025	14:31	LB136789
	Arsenic	5.12	+/-10	U	20.0	P	08/12/2025	14:31	LB136789
	Barium	14.6	+/-50	U	100	P	08/12/2025	14:31	LB136789
	Beryllium	0.56	+/-3	U	6.00	P	08/12/2025	14:31	LB136789
	Cadmium	0.50	+/-3	U	6.00	P	08/12/2025	14:31	LB136789
	Calcium	234	+/-1000	U	2000	P	08/12/2025	14:31	LB136789
	Chromium	2.12	+/-5	U	10.0	P	08/12/2025	14:31	LB136789
	Cobalt	2.26	+/-15	U	30.0	P	08/12/2025	14:31	LB136789
	Copper	4.60	+/-10	U	20.0	P	08/12/2025	14:31	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	14:31	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	14:31	LB136789
	Magnesium	244	+/-1000	U	2000	P	08/12/2025	14:31	LB136789
	Manganese	5.94	+/-10	U	20.0	P	08/12/2025	14:31	LB136789
	Nickel	3.06	+/-20	U	40.0	P	08/12/2025	14:31	LB136789
	Potassium	918	+/-1000	U	2000	P	08/12/2025	14:31	LB136789
	Selenium	9.64	+/-10	U	20.0	P	08/12/2025	14:31	LB136789

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	08/12/2025	14:31	LB136789
	Sodium	868	+/-1000	U	2000	P	08/12/2025	14:31	LB136789
	Thallium	4.38	+/-20	U	40.0	P	08/12/2025	14:31	LB136789
	Vanadium	6.26	+/-20	U	40.0	P	08/12/2025	14:31	LB136789
	Zinc	16.7	+/-20	U	40.0	P	08/12/2025	14:31	LB136789
CCB03	Aluminum	14.5	+/-50	J	100	P	08/12/2025	16:11	LB136789
	Antimony	6.76	+/-25	U	50.0	P	08/12/2025	16:11	LB136789
	Arsenic	5.12	+/-10	U	20.0	P	08/12/2025	16:11	LB136789
	Barium	14.6	+/-50	U	100	P	08/12/2025	16:11	LB136789
	Beryllium	0.56	+/-3	U	6.00	P	08/12/2025	16:11	LB136789
	Cadmium	0.50	+/-3	U	6.00	P	08/12/2025	16:11	LB136789
	Calcium	234	+/-1000	U	2000	P	08/12/2025	16:11	LB136789
	Chromium	2.12	+/-5	U	10.0	P	08/12/2025	16:11	LB136789
	Cobalt	2.26	+/-15	U	30.0	P	08/12/2025	16:11	LB136789
	Copper	4.60	+/-10	U	20.0	P	08/12/2025	16:11	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	16:11	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	16:11	LB136789
	Magnesium	244	+/-1000	U	2000	P	08/12/2025	16:11	LB136789
	Manganese	5.94	+/-10	U	20.0	P	08/12/2025	16:11	LB136789
	Nickel	3.06	+/-20	U	40.0	P	08/12/2025	16:11	LB136789
	Potassium	918	+/-1000	U	2000	P	08/12/2025	16:11	LB136789
	Selenium	9.64	+/-10	U	20.0	P	08/12/2025	16:11	LB136789
	Silver	1.62	+/-5	U	10.0	P	08/12/2025	16:11	LB136789
	Sodium	868	+/-1000	U	2000	P	08/12/2025	16:11	LB136789
	Thallium	4.38	+/-20	U	40.0	P	08/12/2025	16:11	LB136789
	Vanadium	6.26	+/-20	U	40.0	P	08/12/2025	16:11	LB136789
	Zinc	16.7	+/-20	U	40.0	P	08/12/2025	16:11	LB136789
CCB04	Aluminum	22.1	+/-50	J	100	P	08/12/2025	16:34	LB136789
	Antimony	6.76	+/-25	U	50.0	P	08/12/2025	16:34	LB136789
	Arsenic	5.12	+/-10	U	20.0	P	08/12/2025	16:34	LB136789
	Barium	14.6	+/-50	U	100	P	08/12/2025	16:34	LB136789
	Beryllium	0.56	+/-3	U	6.00	P	08/12/2025	16:34	LB136789
	Cadmium	0.50	+/-3	U	6.00	P	08/12/2025	16:34	LB136789
	Calcium	234	+/-1000	U	2000	P	08/12/2025	16:34	LB136789
	Chromium	2.12	+/-5	U	10.0	P	08/12/2025	16:34	LB136789
	Cobalt	2.26	+/-15	U	30.0	P	08/12/2025	16:34	LB136789
	Copper	4.60	+/-10	U	20.0	P	08/12/2025	16:34	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	16:34	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	16:34	LB136789

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	08/12/2025	16:34	LB136789
	Manganese	5.94	+/-10	U	20.0	P	08/12/2025	16:34	LB136789
	Nickel	3.06	+/-20	U	40.0	P	08/12/2025	16:34	LB136789
	Potassium	918	+/-1000	U	2000	P	08/12/2025	16:34	LB136789
	Selenium	9.64	+/-10	U	20.0	P	08/12/2025	16:34	LB136789
	Silver	1.62	+/-5	U	10.0	P	08/12/2025	16:34	LB136789
	Sodium	868	+/-1000	U	2000	P	08/12/2025	16:34	LB136789
	Thallium	4.38	+/-20	U	40.0	P	08/12/2025	16:34	LB136789
	Vanadium	6.26	+/-20	U	40.0	P	08/12/2025	16:34	LB136789
	Zinc	16.7	+/-20	U	40.0	P	08/12/2025	16:34	LB136789

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169219BL		WATER		Batch Number:	PB169219		Prep Date:	08/12/2025	
	Mercury	0.027	<0.2	U	0.20	CV	08/12/2025	17:57	LB136791

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169195BL	WATER			Batch Number:	PB169195		Prep Date:	08/11/2025	
	Aluminum	15.4	<25	U	50.0	P	08/12/2025	12:58	LB136789
	Antimony	2.72	<12.5	U	25.0	P	08/12/2025	12:58	LB136789
	Arsenic	1.85	<5	U	10.0	P	08/12/2025	12:58	LB136789
	Barium	9.75	<25	U	50.0	P	08/12/2025	12:58	LB136789
	Beryllium	0.27	<1.5	U	3.00	P	08/12/2025	12:58	LB136789
	Cadmium	0.31	<1.5	U	3.00	P	08/12/2025	12:58	LB136789
	Calcium	86.4	<500	U	1000	P	08/12/2025	12:58	LB136789
	Chromium	0.53	<2.5	U	5.00	P	08/12/2025	12:58	LB136789
	Cobalt	1.16	<7.5	U	15.0	P	08/12/2025	12:58	LB136789
	Copper	1.89	<5	U	10.0	P	08/12/2025	12:58	LB136789
	Iron	18.0	<25	U	50.0	P	08/12/2025	12:58	LB136789
	Lead	1.21	<3	U	6.00	P	08/12/2025	12:58	LB136789
	Magnesium	104	<500	U	1000	P	08/12/2025	12:58	LB136789
	Manganese	2.46	<5	U	10.0	P	08/12/2025	12:58	LB136789
	Nickel	1.90	<10	U	20.0	P	08/12/2025	12:58	LB136789
	Potassium	309	<500	U	1000	P	08/12/2025	12:58	LB136789
	Selenium	2.52	<5	U	10.0	P	08/12/2025	12:58	LB136789
	Silver	0.81	<2.5	U	5.00	P	08/12/2025	12:58	LB136789
	Sodium	367	<500	U	1000	P	08/12/2025	12:58	LB136789
	Thallium	2.51	<10	U	20.0	P	08/12/2025	12:58	LB136789
	Vanadium	2.49	<10	U	20.0	P	08/12/2025	12:58	LB136789
	Zinc	2.00	<10	U	20.0	P	08/12/2025	12:58	LB136789



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV11	Mercury	4.06	4.0	102	95 - 105	CV	08/12/2025	17:35	LB136791

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV32	Mercury	4.97	5.0	99	90 - 110	CV	08/12/2025	17:42	LB136791
CCV33	Mercury	4.66	5.0	93	90 - 110	CV	08/12/2025	18:18	LB136791
CCV34	Mercury	4.75	5.0	95	90 - 110	CV	08/12/2025	18:30	LB136791

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7680	8000	96	95 - 105	P	08/12/2025	11:34	LB136789
	Antimony	3940	4000	98	95 - 105	P	08/12/2025	11:34	LB136789
	Arsenic	3830	4000	96	95 - 105	P	08/12/2025	11:34	LB136789
	Barium	7750	8000	97	95 - 105	P	08/12/2025	11:34	LB136789
	Beryllium	191	200	96	95 - 105	P	08/12/2025	11:34	LB136789
	Cadmium	1900	2000	95	95 - 105	P	08/12/2025	11:34	LB136789
	Calcium	19300	20000	96	95 - 105	P	08/12/2025	11:34	LB136789
	Chromium	794	800	99	95 - 105	P	08/12/2025	11:34	LB136789
	Cobalt	1950	2000	98	95 - 105	P	08/12/2025	11:34	LB136789
	Copper	999	1000	100	95 - 105	P	08/12/2025	11:34	LB136789
	Iron	3930	4000	98	95 - 105	P	08/12/2025	11:34	LB136789
	Lead	4050	4000	101	95 - 105	P	08/12/2025	11:34	LB136789
	Magnesium	19200	20000	96	95 - 105	P	08/12/2025	11:34	LB136789
	Manganese	1930	2000	97	95 - 105	P	08/12/2025	11:34	LB136789
	Nickel	1950	2000	98	95 - 105	P	08/12/2025	11:34	LB136789
	Potassium	19200	20000	96	95 - 105	P	08/12/2025	11:34	LB136789
	Selenium	3800	4000	95	95 - 105	P	08/12/2025	11:34	LB136789
	Silver	1010	1000	101	95 - 105	P	08/12/2025	11:34	LB136789
	Sodium	19200	20000	96	95 - 105	P	08/12/2025	11:34	LB136789
	Thallium	3860	4000	96	95 - 105	P	08/12/2025	11:34	LB136789
	Vanadium	1940	2000	97	95 - 105	P	08/12/2025	11:34	LB136789
	Zinc	1950	2000	97	95 - 105	P	08/12/2025	11:34	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	99.4	100	99	80 - 120	P	08/12/2025	11:43	LB136789
	Antimony	51.1	50.0	102	80 - 120	P	08/12/2025	11:43	LB136789
	Arsenic	22.2	20.0	111	80 - 120	P	08/12/2025	11:43	LB136789
	Barium	91.1	100	91	80 - 120	P	08/12/2025	11:43	LB136789
	Beryllium	5.84	6.0	97	80 - 120	P	08/12/2025	11:43	LB136789
	Cadmium	5.45	6.0	91	80 - 120	P	08/12/2025	11:43	LB136789
	Calcium	1990	2000	99	80 - 120	P	08/12/2025	11:43	LB136789
	Chromium	9.95	10.0	100	80 - 120	P	08/12/2025	11:43	LB136789
	Cobalt	29.0	30.0	97	80 - 120	P	08/12/2025	11:43	LB136789
	Copper	21.2	20.0	106	80 - 120	P	08/12/2025	11:43	LB136789
	Iron	102	100	102	80 - 120	P	08/12/2025	11:43	LB136789
	Lead	10.9	12.0	91	80 - 120	P	08/12/2025	11:43	LB136789
	Magnesium	2090	2000	104	80 - 120	P	08/12/2025	11:43	LB136789
	Manganese	20.5	20.0	103	80 - 120	P	08/12/2025	11:43	LB136789
	Nickel	39.0	40.0	97	80 - 120	P	08/12/2025	11:43	LB136789
	Potassium	1730	2000	86	80 - 120	P	08/12/2025	11:43	LB136789
	Selenium	18.2	20.0	91	80 - 120	P	08/12/2025	11:43	LB136789
	Silver	10.2	10.0	102	80 - 120	P	08/12/2025	11:43	LB136789
	Sodium	1730	2000	86	80 - 120	P	08/12/2025	11:43	LB136789
	Thallium	33.4	40.0	83	80 - 120	P	08/12/2025	11:43	LB136789
	Vanadium	42.7	40.0	107	80 - 120	P	08/12/2025	11:43	LB136789
	Zinc	40.5	40.0	101	80 - 120	P	08/12/2025	11:43	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9600	10000	96	90 - 110	P	08/12/2025	12:50	LB136789
	Antimony	4970	5000	99	90 - 110	P	08/12/2025	12:50	LB136789
	Arsenic	4910	5000	98	90 - 110	P	08/12/2025	12:50	LB136789
	Barium	9910	10000	99	90 - 110	P	08/12/2025	12:50	LB136789
	Beryllium	249	250	99	90 - 110	P	08/12/2025	12:50	LB136789
	Cadmium	2380	2500	95	90 - 110	P	08/12/2025	12:50	LB136789
	Calcium	24100	25000	96	90 - 110	P	08/12/2025	12:50	LB136789
	Chromium	968	1000	97	90 - 110	P	08/12/2025	12:50	LB136789
	Cobalt	2390	2500	96	90 - 110	P	08/12/2025	12:50	LB136789
	Copper	1220	1250	98	90 - 110	P	08/12/2025	12:50	LB136789
	Iron	5030	5000	101	90 - 110	P	08/12/2025	12:50	LB136789
	Lead	4880	5000	98	90 - 110	P	08/12/2025	12:50	LB136789
	Magnesium	24100	25000	96	90 - 110	P	08/12/2025	12:50	LB136789
	Manganese	2420	2500	97	90 - 110	P	08/12/2025	12:50	LB136789
	Nickel	2400	2500	96	90 - 110	P	08/12/2025	12:50	LB136789
	Potassium	25000	25000	100	90 - 110	P	08/12/2025	12:50	LB136789
	Selenium	4910	5000	98	90 - 110	P	08/12/2025	12:50	LB136789
	Silver	1230	1250	98	90 - 110	P	08/12/2025	12:50	LB136789
	Sodium	25000	25000	100	90 - 110	P	08/12/2025	12:50	LB136789
	Thallium	4870	5000	97	90 - 110	P	08/12/2025	12:50	LB136789
	Vanadium	2440	2500	98	90 - 110	P	08/12/2025	12:50	LB136789
	Zinc	2480	2500	99	90 - 110	P	08/12/2025	12:50	LB136789
CCV02	Aluminum	9870	10000	99	90 - 110	P	08/12/2025	14:27	LB136789
	Antimony	5220	5000	104	90 - 110	P	08/12/2025	14:27	LB136789
	Arsenic	5160	5000	103	90 - 110	P	08/12/2025	14:27	LB136789
	Barium	9880	10000	99	90 - 110	P	08/12/2025	14:27	LB136789
	Beryllium	229	250	91	90 - 110	P	08/12/2025	14:27	LB136789
	Cadmium	2460	2500	98	90 - 110	P	08/12/2025	14:27	LB136789
	Calcium	24300	25000	97	90 - 110	P	08/12/2025	14:27	LB136789
	Chromium	1000	1000	100	90 - 110	P	08/12/2025	14:27	LB136789
	Cobalt	2470	2500	99	90 - 110	P	08/12/2025	14:27	LB136789
	Copper	1270	1250	102	90 - 110	P	08/12/2025	14:27	LB136789
	Iron	5300	5000	106	90 - 110	P	08/12/2025	14:27	LB136789
	Lead	4880	5000	98	90 - 110	P	08/12/2025	14:27	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Environmental Restoration, LLC

Contract: ENVI60

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

SDG No.: Q2799

Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24000	25000	96	90 - 110	P	08/12/2025	14:27	LB136789
	Manganese	2440	2500	97	90 - 110	P	08/12/2025	14:27	LB136789
	Nickel	2480	2500	99	90 - 110	P	08/12/2025	14:27	LB136789
	Potassium	26700	25000	107	90 - 110	P	08/12/2025	14:27	LB136789
	Selenium	5250	5000	105	90 - 110	P	08/12/2025	14:27	LB136789
	Silver	1260	1250	101	90 - 110	P	08/12/2025	14:27	LB136789
	Sodium	24700	25000	99	90 - 110	P	08/12/2025	14:27	LB136789
	Thallium	5280	5000	106	90 - 110	P	08/12/2025	14:27	LB136789
	Vanadium	2490	2500	100	90 - 110	P	08/12/2025	14:27	LB136789
	Zinc	2560	2500	102	90 - 110	P	08/12/2025	14:27	LB136789
CCV03	Aluminum	9920	10000	99	90 - 110	P	08/12/2025	16:07	LB136789
	Antimony	5180	5000	104	90 - 110	P	08/12/2025	16:07	LB136789
	Arsenic	5110	5000	102	90 - 110	P	08/12/2025	16:07	LB136789
	Barium	9960	10000	100	90 - 110	P	08/12/2025	16:07	LB136789
	Beryllium	229	250	91	90 - 110	P	08/12/2025	16:07	LB136789
	Cadmium	2430	2500	97	90 - 110	P	08/12/2025	16:07	LB136789
	Calcium	24200	25000	97	90 - 110	P	08/12/2025	16:07	LB136789
	Chromium	990	1000	99	90 - 110	P	08/12/2025	16:07	LB136789
	Cobalt	2440	2500	98	90 - 110	P	08/12/2025	16:07	LB136789
	Copper	1260	1250	101	90 - 110	P	08/12/2025	16:07	LB136789
	Iron	5310	5000	106	90 - 110	P	08/12/2025	16:07	LB136789
	Lead	4820	5000	96	90 - 110	P	08/12/2025	16:07	LB136789
	Magnesium	23900	25000	96	90 - 110	P	08/12/2025	16:07	LB136789
	Manganese	2440	2500	98	90 - 110	P	08/12/2025	16:07	LB136789
	Nickel	2450	2500	98	90 - 110	P	08/12/2025	16:07	LB136789
	Potassium	26400	25000	106	90 - 110	P	08/12/2025	16:07	LB136789
	Selenium	5200	5000	104	90 - 110	P	08/12/2025	16:07	LB136789
	Silver	1260	1250	100	90 - 110	P	08/12/2025	16:07	LB136789
	Sodium	24900	25000	100	90 - 110	P	08/12/2025	16:07	LB136789
	Thallium	5060	5000	101	90 - 110	P	08/12/2025	16:07	LB136789
	Vanadium	2490	2500	100	90 - 110	P	08/12/2025	16:07	LB136789
	Zinc	2520	2500	101	90 - 110	P	08/12/2025	16:07	LB136789
CCV04	Aluminum	9960	10000	100	90 - 110	P	08/12/2025	16:30	LB136789
	Antimony	5290	5000	106	90 - 110	P	08/12/2025	16:30	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5230	5000	104	90 - 110	P	08/12/2025	16:30	LB136789
	Barium	10100	10000	101	90 - 110	P	08/12/2025	16:30	LB136789
	Beryllium	227	250	91	90 - 110	P	08/12/2025	16:30	LB136789
	Cadmium	2460	2500	98	90 - 110	P	08/12/2025	16:30	LB136789
	Calcium	24200	25000	97	90 - 110	P	08/12/2025	16:30	LB136789
	Chromium	995	1000	100	90 - 110	P	08/12/2025	16:30	LB136789
	Cobalt	2470	2500	99	90 - 110	P	08/12/2025	16:30	LB136789
	Copper	1280	1250	103	90 - 110	P	08/12/2025	16:30	LB136789
	Iron	5400	5000	108	90 - 110	P	08/12/2025	16:30	LB136789
	Lead	4870	5000	97	90 - 110	P	08/12/2025	16:30	LB136789
	Magnesium	23900	25000	96	90 - 110	P	08/12/2025	16:30	LB136789
	Manganese	2440	2500	98	90 - 110	P	08/12/2025	16:30	LB136789
	Nickel	2490	2500	100	90 - 110	P	08/12/2025	16:30	LB136789
	Potassium	26700	25000	107	90 - 110	P	08/12/2025	16:30	LB136789
	Selenium	5330	5000	107	90 - 110	P	08/12/2025	16:30	LB136789
	Silver	1270	1250	101	90 - 110	P	08/12/2025	16:30	LB136789
	Sodium	25200	25000	101	90 - 110	P	08/12/2025	16:30	LB136789
	Thallium	5020	5000	100	90 - 110	P	08/12/2025	16:30	LB136789
	Vanadium	2500	2500	100	90 - 110	P	08/12/2025	16:30	LB136789
	Zinc	2500	2500	100	90 - 110	P	08/12/2025	16:30	LB136789

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	103	100	103	65 - 135	P	08/12/2025	11:56	LB136789
	Antimony	48.6	50.0	97	65 - 135	P	08/12/2025	11:56	LB136789
	Arsenic	21.6	20.0	108	65 - 135	P	08/12/2025	11:56	LB136789
	Barium	91.6	100	92	65 - 135	P	08/12/2025	11:56	LB136789
	Beryllium	5.75	6.0	96	65 - 135	P	08/12/2025	11:56	LB136789
	Cadmium	5.34	6.0	89	65 - 135	P	08/12/2025	11:56	LB136789
	Calcium	1970	2000	98	65 - 135	P	08/12/2025	11:56	LB136789
	Chromium	9.88	10.0	99	65 - 135	P	08/12/2025	11:56	LB136789
	Cobalt	28.5	30.0	95	65 - 135	P	08/12/2025	11:56	LB136789
	Copper	20.3	20.0	102	65 - 135	P	08/12/2025	11:56	LB136789
	Iron	98.1	100	98	65 - 135	P	08/12/2025	11:56	LB136789
	Lead	11.1	12.0	93	65 - 135	P	08/12/2025	11:56	LB136789
	Magnesium	2060	2000	103	65 - 135	P	08/12/2025	11:56	LB136789
	Manganese	20.4	20.0	102	65 - 135	P	08/12/2025	11:56	LB136789
	Nickel	38.5	40.0	96	65 - 135	P	08/12/2025	11:56	LB136789
	Potassium	1640	2000	82	65 - 135	P	08/12/2025	11:56	LB136789
	Selenium	17.3	20.0	86	65 - 135	P	08/12/2025	11:56	LB136789
	Silver	10.4	10.0	104	65 - 135	P	08/12/2025	11:56	LB136789
	Sodium	1670	2000	84	65 - 135	P	08/12/2025	11:56	LB136789
	Thallium	32.8	40.0	82	65 - 135	P	08/12/2025	11:56	LB136789
	Vanadium	39.2	40.0	98	65 - 135	P	08/12/2025	11:56	LB136789
	Zinc	40.0	40.0	100	65 - 135	P	08/12/2025	11:56	LB136789
CRA	Mercury	0.038	0.05	76	60 - 140	CV	08/12/2025	17:47	LB136791

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Environmental Restoration, LLC</u>	SDG No.: <u>Q2799</u>
Contract: <u>ENVI60</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	237000	255000	93	216000	294000	08/12/2025	12:00	LB136789
	Antimony	-4.63			-50	50	08/12/2025	12:00	LB136789
	Arsenic	9.40			-20	20	08/12/2025	12:00	LB136789
	Barium	-3.44	6.0	57	-94	106	08/12/2025	12:00	LB136789
	Beryllium	1.34			-6	6	08/12/2025	12:00	LB136789
	Cadmium	1.01	1.0	101	-5	7	08/12/2025	12:00	LB136789
	Calcium	227000	245000	93	208000	282000	08/12/2025	12:00	LB136789
	Chromium	57.4	52.0	110	42	62	08/12/2025	12:00	LB136789
	Cobalt	1.47			-30	30	08/12/2025	12:00	LB136789
	Copper	4.88	2.0	244	-18	22	08/12/2025	12:00	LB136789
	Iron	98300	101000	97	85600	116500	08/12/2025	12:00	LB136789
	Lead	-7.54			-12	12	08/12/2025	12:00	LB136789
	Magnesium	240000	255000	94	216000	294000	08/12/2025	12:00	LB136789
	Manganese	7.76	7.0	111	-13	27	08/12/2025	12:00	LB136789
	Nickel	5.34	2.0	267	-38	42	08/12/2025	12:00	LB136789
	Potassium	-88.4			0	0	08/12/2025	12:00	LB136789
	Selenium	3.67			-20	20	08/12/2025	12:00	LB136789
	Silver	2.23			-10	10	08/12/2025	12:00	LB136789
	Sodium	-85.2			0	0	08/12/2025	12:00	LB136789
	Thallium	1.31			-40	40	08/12/2025	12:00	LB136789
	Vanadium	3.28			-40	40	08/12/2025	12:00	LB136789
	Zinc	2.31			-40	40	08/12/2025	12:00	LB136789
ICSAB01	Aluminum	242000	247000	98	209000	285000	08/12/2025	12:06	LB136789
	Antimony	584	618	94	525	711	08/12/2025	12:06	LB136789
	Arsenic	109	104	105	88.4	120	08/12/2025	12:06	LB136789
	Barium	483	537	90	437	637	08/12/2025	12:06	LB136789
	Beryllium	474	495	96	420	570	08/12/2025	12:06	LB136789
	Cadmium	957	972	98	826	1120	08/12/2025	12:06	LB136789
	Calcium	231000	235000	98	199000	271000	08/12/2025	12:06	LB136789
	Chromium	547	542	101	460	624	08/12/2025	12:06	LB136789
	Cobalt	484	476	102	404	548	08/12/2025	12:06	LB136789
	Copper	475	511	93	434	588	08/12/2025	12:06	LB136789
	Iron	100000	99300	101	84400	114500	08/12/2025	12:06	LB136789
	Lead	39.5	49.0	81	37	61	08/12/2025	12:06	LB136789
	Magnesium	243000	248000	98	210000	286000	08/12/2025	12:06	LB136789
	Manganese	483	507	95	430	584	08/12/2025	12:06	LB136789
	Nickel	966	954	101	810	1100	08/12/2025	12:06	LB136789
	Potassium	-167			0	0	08/12/2025	12:06	LB136789
	Selenium	53.2	46.0	116	26	66	08/12/2025	12:06	LB136789
	Silver	212	201	106	170	232	08/12/2025	12:06	LB136789
	Sodium	-77.7			0	0	08/12/2025	12:06	LB136789
	Thallium	98.4	108	91	68	148	08/12/2025	12:06	LB136789

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Environmental Restoration, LLC</u>	SDG No.: <u>Q2799</u>
Contract: <u>ENVI60</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	473	491	96	417	565	08/12/2025	12:06	LB136789
	Zinc	1010	952	106	809	1095	08/12/2025	12:06	LB136789
ICSA	Aluminum	233000	255000	91	216000	294000	08/12/2025	12:10	LB136789
	Antimony	-41.0			-50	50	08/12/2025	12:10	LB136789
	Arsenic	43.5			-20	20	08/12/2025	12:10	LB136789
	Barium	-107	6.0	1783	-94	106	08/12/2025	12:10	LB136789
	Beryllium	2.59			-6	6	08/12/2025	12:10	LB136789
	Cadmium	-2.80	1.0	280	-5	7	08/12/2025	12:10	LB136789
	Calcium	233000	245000	95	208000	282000	08/12/2025	12:10	LB136789
	Chromium	57.2	52.0	110	42	62	08/12/2025	12:10	LB136789
	Cobalt	-4.02			-30	30	08/12/2025	12:10	LB136789
	Copper	4.26	2.0	213	-18	22	08/12/2025	12:10	LB136789
	Iron	94000	101000	93	85600	116500	08/12/2025	12:10	LB136789
	Lead	-5.72			-12	12	08/12/2025	12:10	LB136789
	Magnesium	241000	255000	94	216000	294000	08/12/2025	12:10	LB136789
	Manganese	2.24	7.0	32	-13	27	08/12/2025	12:10	LB136789
	Nickel	0.64	2.0	32	-38	42	08/12/2025	12:10	LB136789
	Potassium	-5920			0	0	08/12/2025	12:10	LB136789
	Selenium	-20.3			-20	20	08/12/2025	12:10	LB136789
	Silver	5.47			-10	10	08/12/2025	12:10	LB136789
	Sodium	-4070			0	0	08/12/2025	12:10	LB136789
	Thallium	-35.9			-40	40	08/12/2025	12:10	LB136789
	Vanadium	35.3			-40	40	08/12/2025	12:10	LB136789
	Zinc	3.13			-40	40	08/12/2025	12:10	LB136789
ICSAB	Aluminum	234000	247000	95	209000	285000	08/12/2025	12:15	LB136789
	Antimony	586	618	95	525	711	08/12/2025	12:15	LB136789
	Arsenic	110	104	106	88.4	120	08/12/2025	12:15	LB136789
	Barium	358	537	67	437	637	08/12/2025	12:15	LB136789
	Beryllium	477	495	96	420	570	08/12/2025	12:15	LB136789
	Cadmium	944	972	97	826	1120	08/12/2025	12:15	LB136789
	Calcium	234000	235000	100	199000	271000	08/12/2025	12:15	LB136789
	Chromium	553	542	102	460	624	08/12/2025	12:15	LB136789
	Cobalt	477	476	100	404	548	08/12/2025	12:15	LB136789
	Copper	510	511	100	434	588	08/12/2025	12:15	LB136789
	Iron	95800	99300	96	84400	114500	08/12/2025	12:15	LB136789
	Lead	17.4	49.0	36	37	61	08/12/2025	12:15	LB136789
	Magnesium	242000	248000	98	210000	286000	08/12/2025	12:15	LB136789
	Manganese	480	507	95	430	584	08/12/2025	12:15	LB136789
	Nickel	962	954	101	810	1100	08/12/2025	12:15	LB136789
	Potassium	-6140			0	0	08/12/2025	12:15	LB136789
	Selenium	11.6	46.0	25	26	66	08/12/2025	12:15	LB136789
	Silver	198	201	98	170	232	08/12/2025	12:15	LB136789

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Environmental Restoration, LLC</u>	SDG No.: <u>Q2799</u>
Contract: <u>ENVI60</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Sodium	-3980			0	0	08/12/2025	12:15	LB136789
	Thallium	89.3	108	83	68	148	08/12/2025	12:15	LB136789
	Vanadium	552	491	112	417	565	08/12/2025	12:15	LB136789
	Zinc	975	952	102	809	1095	08/12/2025	12:15	LB136789



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Environmental Restoration, LLC **level:** low **sdg no.:** Q2799
contract: ENVI60 **lab code:** ACE
matrix: Water **sample id:** Q2785-01 **client id:** WATERMS
Percent Solids for Sample: NA **Spiked ID:** Q2785-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.83		0.20	U	4.0	96		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Environmental Restoration, LLC **level:** low **sdg no.:** Q2799
contract: ENVI60 **lab code:** ACE
matrix: Water **sample id:** Q2785-01 **client id:** WATERMSD
Percent Solids for Sample: NA **Spiked ID:** Q2785-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.83		0.20	U	4.0	96		CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Environmental Restoration, LLC **level:** low **sdg no.:** Q2799
contract: ENVI60 **lab code:** ACE
matrix: Water **sample id:** Q2811-02 **client id:** EFF-WASTE WATERMS
Percent Solids for Sample: NA **Spiked ID:** Q2811-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	976		50.3		1000	93		P
Antimony	ug/L	75 - 125	401		25.0	U	400	100		P
Arsenic	ug/L	75 - 125	399		10.0	U	400	100		P
Barium	ug/L	75 - 125	95.9		50.0	U	100	96		P
Beryllium	ug/L	75 - 125	83.6		3.00	U	100	84		P
Cadmium	ug/L	75 - 125	93.8		3.00	U	100	94		P
Calcium	ug/L	75 - 125	13000		12400		500	111		P
Chromium	ug/L	75 - 125	194		0.65	J	200	96		P
Cobalt	ug/L	75 - 125	94.8		15.0	U	100	95		P
Copper	ug/L	75 - 125	223		84.6		150	92		P
Iron	ug/L	75 - 125	1820		201		1500	108		P
Lead	ug/L	75 - 125	430		1.76	J	500	86		P
Magnesium	ug/L	75 - 125	3890		3000		1000	89		P
Manganese	ug/L	75 - 125	147		55.3		100	92		P
Nickel	ug/L	75 - 125	237		20.0	U	250	95		P
Potassium	ug/L	75 - 125	8360		2190		5000	123		P
Selenium	ug/L	75 - 125	999		3.58	J	1000	100		P
Silver	ug/L	75 - 125	31.1		5.00	U	37.5	83		P
Sodium	ug/L	75 - 125	530000		534000		1500	-222		P
Thallium	ug/L	75 - 125	759		20.0	U	1000	76		P
Vanadium	ug/L	75 - 125	143		20.0	U	150	95		P
Zinc	ug/L	75 - 125	303		203		100	99		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Environmental Restoration, LLC **level:** low **sdg no.:** Q2799
contract: ENVI60 **lab code:** ACE
matrix: Water **sample id:** Q2811-02 **client id:** EFF-WASTE WATERMSD
Percent Solids for Sample: NA **Spiked ID:** Q2811-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	1040		50.3		1000	99		P
Antimony	ug/L	75 - 125	432		25.0	U	400	108		P
Arsenic	ug/L	75 - 125	438		10.0	U	400	110		P
Barium	ug/L	75 - 125	101		50.0	U	100	101		P
Beryllium	ug/L	75 - 125	88.5		3.00	U	100	88		P
Cadmium	ug/L	75 - 125	100		3.00	U	100	100		P
Calcium	ug/L	75 - 125	14300		12400		500	380		P
Chromium	ug/L	75 - 125	203		0.65	J	200	101		P
Cobalt	ug/L	75 - 125	101		15.0	U	100	101		P
Copper	ug/L	75 - 125	243		84.6		150	106		P
Iron	ug/L	75 - 125	1970		201		1500	118		P
Lead	ug/L	75 - 125	465		1.76	J	500	93		P
Magnesium	ug/L	75 - 125	4260		3000		1000	126	N	P
Manganese	ug/L	75 - 125	157		55.3		100	102		P
Nickel	ug/L	75 - 125	253		20.0	U	250	101		P
Potassium	ug/L	75 - 125	9530		2190		5000	147	N	P
Selenium	ug/L	75 - 125	1100		3.58	J	1000	110		P
Silver	ug/L	75 - 125	34.9		5.00	U	37.5	93		P
Sodium	ug/L	75 - 125	584000		534000		1500	3361		P
Thallium	ug/L	75 - 125	925		20.0	U	1000	92		P
Vanadium	ug/L	75 - 125	151		20.0	U	150	101		P
Zinc	ug/L	75 - 125	342		203		100	138	N	P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: <u>Environmental Restoration, LLC</u>	SDG No.: <u>Q2799</u>	
Contract: <u>ENVI60</u>	Lab Code: <u>ACE</u>	
Matrix: <u>Water</u>	Level: <u>LOW</u>	Client ID: <u>EFF-WASTE WATERA</u>
Sample ID: <u>Q2811-02</u>	Spiked ID: <u>Q2811-02A</u>	

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Magnesium	ug/L	85 - 115	3870		3000		1000	86		P
Potassium	ug/L	85 - 115	8330		2190		5000	123	N	P
Zinc	ug/L	85 - 115	305		203		100	102		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Environmental Restoration, LLC **Level:** LOW **SDG No.:** Q2799
Contract: ENVI60 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2785-01 **Client ID:** WATERDUP
Percent Solids for Sample: NA **Duplicate ID** Q2785-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Environmental Restoration, LLC</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2799</u>
Contract:	<u>ENVI60</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2785-01MS</u>	Client ID:	<u>WATERMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2785-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.83		3.83		0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Environmental Restoration, LLC **Level:** LOW **SDG No.:** Q2799
Contract: ENVI60 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2811-02 **Client ID:** EFF-WASTE WATERDUP
Percent Solids for Sample: NA **Duplicate ID** Q2811-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	50.3		49.6	J	1		P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	50.0	U	50.0	U			P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Calcium	ug/L	20	12400		12300		1		P
Chromium	ug/L	20	0.65	J	0.62	J	6		P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	84.6		83.4		1		P
Iron	ug/L	20	201		191		5		P
Lead	ug/L	20	1.76	J	1.22	J	36		P
Magnesium	ug/L	20	3000		2970		1		P
Manganese	ug/L	20	55.3		53.9		3		P
Nickel	ug/L	20	20.0	U	20.0	U			P
Potassium	ug/L	20	2190		2250		3		P
Selenium	ug/L	20	3.58	J	10.0	U	200.0		P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	534000		514000		4		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	203		199		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Environmental Restoration, LLC **Level:** LOW **SDG No.:** Q2799
Contract: ENVI60 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2811-02MS **Client ID:** EFF-WASTE WATERMSD
Percent Solids for Sample: NA **Duplicate ID** Q2811-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	976		1040		6		P
Antimony	ug/L	20	401		432		7		P
Arsenic	ug/L	20	399		438		9		P
Barium	ug/L	20	95.9		101		5		P
Beryllium	ug/L	20	83.6		88.5		6		P
Cadmium	ug/L	20	93.8		100		6		P
Calcium	ug/L	20	13000		14300		10		P
Chromium	ug/L	20	194		203		5		P
Cobalt	ug/L	20	94.8		101		6		P
Copper	ug/L	20	223		243		9		P
Iron	ug/L	20	1820		1970		8		P
Lead	ug/L	20	430		465		8		P
Magnesium	ug/L	20	3890		4260		9		P
Manganese	ug/L	20	147		157		7		P
Nickel	ug/L	20	237		253		7		P
Potassium	ug/L	20	8360		9530		13		P
Selenium	ug/L	20	999		1100		10		P
Silver	ug/L	20	31.1		34.9		12		P
Sodium	ug/L	20	530000		584000		10		P
Thallium	ug/L	20	759		925		20		P
Vanadium	ug/L	20	143		151		5		P
Zinc	ug/L	20	303		342		12		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169195BS							
Aluminum	ug/L	1000	963		96	85 - 115	P
Antimony	ug/L	400	400		100	85 - 115	P
Arsenic	ug/L	400	383		96	85 - 115	P
Barium	ug/L	100	92.9		93	85 - 115	P
Beryllium	ug/L	100	93.7		94	85 - 115	P
Cadmium	ug/L	100	92.5		92	85 - 115	P
Calcium	ug/L	500	486	J	97	85 - 115	P
Chromium	ug/L	200	197		98	85 - 115	P
Cobalt	ug/L	100	95.4		95	85 - 115	P
Copper	ug/L	150	151		101	85 - 115	P
Iron	ug/L	1500	1510		101	85 - 115	P
Lead	ug/L	500	463		93	85 - 115	P
Magnesium	ug/L	1000	961	J	96	85 - 115	P
Manganese	ug/L	100	99.0		99	85 - 115	P
Nickel	ug/L	250	240		96	85 - 115	P
Potassium	ug/L	5000	4740		95	85 - 115	P
Selenium	ug/L	1000	959		96	85 - 115	P
Silver	ug/L	37.5	36.8		98	85 - 115	P
Sodium	ug/L	1500	1330		89	85 - 115	P
Thallium	ug/L	1000	957		96	85 - 115	P
Vanadium	ug/L	150	148		99	85 - 115	P
Zinc	ug/L	100	99.8		100	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169219BS Mercury	ug/L	4.0	3.91		98	85 - 115	CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WATERL

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **Lb No.:** lb136791 **Lab Sample ID :** Q2785-01L **SDG No.:** Q2799
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)	Serial Dilution Result (S)	% Differ- ence	Q	M
Mercury	0.20 U	1.00 U			CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

EFF-WASTE WATERL

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **Lb No.:** lb136789 **Lab Sample ID :** Q2811-02L **SDG No.:** Q2799
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ-ence	Q	M
		C		C			
Aluminum	50.3		250	U	100.0		P
Antimony	25.0	U	125	U			P
Arsenic	10.0	U	50.0	U			P
Barium	50.0	U	250	U			P
Beryllium	3.00	U	15.0	U			P
Cadmium	3.00	U	15.0	U			P
Calcium	12400		13300		7		P
Chromium	0.65	J	25.0	U	100.0		P
Cobalt	15.0	U	75.0	U			P
Copper	84.6		86.8		3		P
Iron	201		194	J	3		P
Lead	1.76	J	30.0	U	100.0		P
Magnesium	3000		3210	J	7		P
Manganese	55.3		57.8		5		P
Nickel	20.0	U	100	U			P
Potassium	2190		8010		266		P
Selenium	3.58	J	50.0	U	100.0		P
Silver	5.00	U	25.0	U			P
Sodium	534000		469000		12		P
Thallium	20.0	U	100	U			P
Vanadium	20.0	U	100	U			P
Zinc	203		199		2		P

metals
- 14 -
ANALYSIS RUN LOG

Client: Environmental Restoration, LLC

Contract: ENVI60

Lab code: ACE

Sdg no.: Q2799

Instrument id number: **Method:**

Run number: LB136789

Start date: 08/12/2025 **End date:** 08/12/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1039	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1044	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1048	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1052	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1056	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1100	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1134	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1143	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1148	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1156	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1200	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1206	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA	ICSA	20	1210	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB	ICSAB	20	1215	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1250	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1254	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB169195BL	PB169195BL	1	1258	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB169195BS	PB169195BS	1	1308	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2799-01	CC-071325-RW	1	1328	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2811-02DUP	EFF-WASTE WATERDUP	1	1337	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2811-02L	EFF-WASTE WATERL	5	1341	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2811-02MS	EFF-WASTE WATERMS	1	1346	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2811-02MSD	EFF-WASTE WATERMSD	1	1350	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1427	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1431	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2811-02A	EFF-WASTE WATERA	1	1435	K,Mg,Zn
CCV03	CCV03	1	1607	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1611	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1630	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1634	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Environmental Restoration, LLC

Contract: ENVI60

Lab code: ACE

Sdg no.: Q2799

Instrument id number: _____ **Method:** _____

Run number: LB136791

Start date: 08/12/2025 **End date:** 08/12/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1658	HG
S0.05	S0.05	1	1700	HG
S0.2	S0.2	1	1702	HG
S2.5	S2.5	1	1705	HG
S5	S5	1	1707	HG
S7.5	S7.5	1	1709	HG
S10	S10	1	1728	HG
ICV11	ICV11	1	1735	HG
ICB11	ICB11	1	1740	HG
CCV32	CCV32	1	1742	HG
CCB32	CCB32	1	1745	HG
CRA	CRA	1	1747	HG
PB169219BL	PB169219BL	1	1757	HG
PB169219BS	PB169219BS	1	1759	HG
Q2785-01DUP	WATERDUP	1	1803	HG
Q2785-01MS	WATERMS	1	1811	HG
Q2785-01MSD	WATERMSD	1	1813	HG
CCV33	CCV33	1	1818	HG
CCB33	CCB33	1	1820	HG
Q2799-01	CC-071325-RW	1	1823	HG
Q2785-01L	WATERL	5	1825	HG
CCV34	CCV34	1	1830	HG
CCB34	CCB34	1	1832	HG



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Aluminum	396.100	0.0000000	-0.0002060	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	-0.0075970	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		As	Ba	Be	Cd	Co
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0054900
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Aluminum	396.100	0.0000000	0.0000000	0.0000590	0.0000000	0.0396900
Antimony	206.833	0.0122000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	-0.0000710	-0.0003400
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0017400	-0.0100400
Vanadium	292.402	-0.0025100	0.0000000	0.0000000	0.0000000	-0.0072000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Aluminum	396.100	0.0000000	0.0000000	0.0012800	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	-0.0035600	-0.0007970	0.0000000	-0.0018900	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	-0.0039700	0.0000000	-0.0115600	0.0000000
Vanadium	292.402	0.0000000	0.0005320	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

LAB CHRONICLE

OrderID:	Q2799	OrderDate:	8/7/2025 2:17:00 PM
Client:	Environmental Restoration, LLC	Project:	Cooper Chemical - Long Valley NJ 2-COOP-ANS
Contact:	Byron Hartman	Location:	D21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2799-01	CC-071325-RW	Water			07/14/25			07/14/25
			Mercury	245.1		08/12/25	08/12/25	
			Metals ICP-TAL	200.7		08/11/25	08/12/25	



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169195							
PB169195BL	PB169195BL	MB	WATER	08/11/2025	50.0	25.0	
PB169195BS	PB169195BS	LCS	WATER	08/11/2025	50.0	25.0	
Q2799-01	CC-071325-RW	SAM	WATER	08/11/2025	50.0	25.0	
Q2811-02DUP	EFF-WASTE WATERDUP	DUP	WATER	08/11/2025	50.0	25.0	
Q2811-02MS	EFF-WASTE WATERMS	MS	WATER	08/11/2025	50.0	25.0	
Q2811-02MSD	EFF-WASTE WATERMSD	MSD	WATER	08/11/2025	50.0	25.0	

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2799

Contract: ENVI60

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169219							
PB169219BL	PB169219BL	MB	WATER	08/12/2025	30.0	30.0	
PB169219BS	PB169219BS	LCS	WATER	08/12/2025	30.0	30.0	
Q2785-01DUP	WATERDUP	DUP	WATER	08/12/2025	30.0	30.0	
Q2785-01MS	WATERMS	MS	WATER	08/12/2025	30.0	30.0	
Q2785-01MSD	WATERMSD	MSD	WATER	08/12/2025	30.0	30.0	
Q2799-01	CC-071325-RW	SAM	WATER	08/12/2025	30.0	30.0	

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136789

Review By	Janvi	Review On	8/13/2025 10:20:36 AM
Supervise By	jaswal	Supervise On	8/13/2025 11:12:42 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641		
ICV Standard	MP86752,MP86656		
CCV Standard	MP86646		
ICSA Standard	MP86651,MP86753		
CRI Standard	MP86656		
LCS Standard			
Chk Standard	MP86647,MP86648		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/12/25 10:39		Jaswal	OK
2	S1	S1	CAL2	08/12/25 10:44		Jaswal	OK
3	S2	S2	CAL3	08/12/25 10:48		Jaswal	OK
4	S3	S3	CAL4	08/12/25 10:52		Jaswal	OK
5	S4	S4	CAL5	08/12/25 10:56		Jaswal	OK
6	S5	S5	CAL6	08/12/25 11:00		Jaswal	OK
7	ICV01	ICV01	ICV	08/12/25 11:34		Jaswal	OK
8	LLICV01	LLICV01	LLICV	08/12/25 11:43		Jaswal	OK
9	ICB01	ICB01	ICB	08/12/25 11:48		Jaswal	OK
10	CRI01	CRI01	CRDL	08/12/25 11:56		Jaswal	OK
11	ICSA01	ICSA01	ICSA	08/12/25 12:00		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	08/12/25 12:06		Jaswal	OK
13	ICSADL	ICSADL	ICSA	08/12/25 12:10		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	08/12/25 12:15		Jaswal	OK
15	CCV01	CCV01	CCV	08/12/25 12:50		Jaswal	OK
16	CCB01	CCB01	CCB	08/12/25 12:54		Jaswal	OK
17	PB169195BL	PB169195BL	MB	08/12/25 12:58		Jaswal	OK
18	PB169195BS	PB169195BS	LCS	08/12/25 13:08		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136789

Review By	Janvi	Review On	8/13/2025 10:20:36 AM
Supervise By	jaswal	Supervise On	8/13/2025 11:12:42 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641		
ICV Standard	MP86752,MP86656		
CCV Standard	MP86646		
ICSA Standard	MP86651,MP86753		
CRI Standard	MP86656		
LCS Standard			
Chk Standard	MP86647,MP86648		

19	Q2785-01	WATER	SAM	08/12/25 13:20		Jaswal	OK
20	Q2789-02	Comp	SAM	08/12/25 13:24		Jaswal	OK
21	Q2799-01	CC-071325-RW	SAM	08/12/25 13:28		Jaswal	OK
22	Q2811-02	EFF-WASTE WATER	SAM	08/12/25 13:32		Jaswal	OK
23	Q2811-02DUP	EFF-WASTE WATER	DUP	08/12/25 13:37		Jaswal	OK
24	Q2811-02L	EFF-WASTE WATER	SD	08/12/25 13:41		Jaswal	OK
25	Q2811-02MS	EFF-WASTE WATER	MS	08/12/25 13:46		Jaswal	OK
26	Q2811-02MSD	EFF-WASTE WATER	MSD	08/12/25 13:50		Jaswal	OK
27	CCV02	CCV02	CCV	08/12/25 14:27		Jaswal	OK
28	CCB02	CCB02	CCB	08/12/25 14:31		Jaswal	OK
29	Q2811-02A	EFF-WASTE WATER	PS	08/12/25 14:35	0.1 ML OF EACH M6004 AND M6017	Jaswal	OK
30	Q2798-04	TP-6	SAM	08/12/25 14:40	NOT USE	Jaswal	Not Ok
31	Q2809-01	OR-03-08082025	SAM	08/12/25 15:50		Jaswal	OK
32	Q2828-01	POWDER	SAM	08/12/25 15:54	NOT USE	Jaswal	Not Ok
33	Q2828-01DUP	POWDERDUP	DUP	08/12/25 15:58	NOT USE	Jaswal	Not Ok
34	Q2828-01L	POWDERL	SD	08/12/25 16:03	NOT USE	Jaswal	Not Ok
35	CCV03	CCV03	CCV	08/12/25 16:07		Jaswal	OK
36	CCB03	CCB03	CCB	08/12/25 16:11		Jaswal	OK
37	Q2828-01MS	POWDERMS	MS	08/12/25 16:16	NOT USE	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136789

Review By	Janvi	Review On	8/13/2025 10:20:36 AM
Supervise By	jaswal	Supervise On	8/13/2025 11:12:42 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641		
ICV Standard	MP86752,MP86656		
CCV Standard	MP86646		
ICSA Standard	MP86651,MP86753		
CRI Standard	MP86656		
LCS Standard			
Chk Standard	MP86647,MP86648		

38	Q2828-01MSD	POWDERMSD	MSD	08/12/25 16:21	NOT USE	Jaswal	Not Ok
39	Q2828-01A	POWDERA	PS	08/12/25 16:25	NOT USE,0.1 ML OF EACH M6004 AND M6017	Jaswal	Not Ok
40	CCV04	CCV04	CCV	08/12/25 16:30		Jaswal	OK
41	CCB04	CCB04	CCB	08/12/25 16:34		Jaswal	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136791

Review By	sagar	Review On	8/12/2025 7:03:37 PM
Supervise By	jaswal	Supervise On	8/12/2025 11:49:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86723,MP86723,MP86738,MP86724,MP86725,MP86726,MP86727,MP86729		
ICV Standard	MP86729		
CCV Standard	MP86731		
ICSA Standard			
CRI Standard	MP86733		
LCS Standard			
Chk Standard	MP86730,MP86732,MP86734,MP86737		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/12/25 16:58		sagar	OK
2	S0.05	S0.05	CAL2	08/12/25 17:00		sagar	OK
3	S0.2	S0.2	CAL3	08/12/25 17:02		sagar	OK
4	S2.5	S2.5	CAL4	08/12/25 17:05		sagar	OK
5	S5	S5	CAL5	08/12/25 17:07		sagar	OK
6	S7.5	S7.5	CAL6	08/12/25 17:09		sagar	OK
7	S10	S10	CAL7	08/12/25 17:28		sagar	OK
8	ICV11	ICV11	ICV	08/12/25 17:35		sagar	OK
9	ICB11	ICB11	ICB	08/12/25 17:40		sagar	OK
10	CCV32	CCV32	CCV	08/12/25 17:42		sagar	OK
11	CCB32	CCB32	CCB	08/12/25 17:45		sagar	OK
12	CRA	CRA	CRDL	08/12/25 17:47		sagar	OK
13	HighStd	HighStd	HIGH STD	08/12/25 17:49		sagar	OK
14	ChkStd	ChkStd	SAM	08/12/25 17:51		sagar	OK
15	PB169219BL	PB169219BL	MB	08/12/25 17:57		sagar	OK
16	PB169219BS	PB169219BS	LCS	08/12/25 17:59		sagar	OK
17	Q2785-01	WATER	SAM	08/12/25 18:01		sagar	OK
18	Q2785-01DUP	WATERDUP	DUP	08/12/25 18:03		sagar	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136791

Review By	sagar	Review On	8/12/2025 7:03:37 PM
Supervise By	jaswal	Supervise On	8/12/2025 11:49:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86723,MP86723,MP86738,MP86724,MP86725,MP86726,MP86727,MP86729		
ICV Standard	MP86729		
CCV Standard	MP86731		
ICSA Standard			
CRI Standard	MP86733		
LCS Standard			
Chk Standard	MP86730,MP86732,MP86734,MP86737		

19	Q2785-01MS	WATERMS	MS	08/12/25 18:11		sagar	OK
20	Q2785-01MSD	WATERMSD	MSD	08/12/25 18:13		sagar	OK
21	Q2789-02	Comp	SAM	08/12/25 18:16		sagar	OK
22	CCV33	CCV33	CCV	08/12/25 18:18		sagar	OK
23	CCB33	CCB33	CCB	08/12/25 18:20		sagar	OK
24	Q2799-01	CC-071325-RW	SAM	08/12/25 18:23		sagar	OK
25	Q2785-01L	WATERL	SD	08/12/25 18:25		sagar	OK
26	Q2785-01A	WATERA	PS	08/12/25 18:27		sagar	OK
27	CCV34	CCV34	CCV	08/12/25 18:30		sagar	OK
28	CCB34	CCB34	CCB	08/12/25 18:32		sagar	OK

SOP ID : M200.7-Trace Elements-22

SDG No : N/A

Matrix : WATER

Pipette ID : ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 08/11/2025 Time : 12:55 Temp : 96 °C

End Digest Date: 08/11/2025 Time : 15:55 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: *SL*

Supervisor Signature: *jepp*

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6180
LFS-2	0.25	M6181
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP85156
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/11/25 16:55	<i>SL met dig</i>	<i>jepp / Met dig</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB169195BL	PBW195	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
PB169195BS	LCS195	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	2
Q2769-01	001-WILLETS-PT-BLVD(AUG)	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
Q2769-04	002-35TH-AVE(AUG)	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
Q2785-01	WATER	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	5
Q2789-02	COMP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	6
Q2799-01	CC-071325-RW	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	7
Q2811-02	EFF-WASTE WATER	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	8
Q2811-02MS	EFF-WASTE WATERMS	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	10
Q2811-02MSD	EFF-WASTE WATERMSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	11
Q2811-02DUP	EFF-WASTE WATERDUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	9

SOP ID : M245.1-Mercury-20

SDG No : N/A

Matrix : WATER

Pipette ID: HG A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #1

Block ID: 1. HG HOT BLOCK#1 2. N/A

Start Digest Date: 08/12/2025 Time : 13:50 Temp : 95 °C

End Digest Date: 08/12/2025 Time : 15:55 Temp : 95 °C

Digestion tube ID: M5595

Block thermometer ID: HG-DIG#1

Dig Technician Signature: *[Signature]*

Supervisor Signature: *[Signature]*

Temp : 1. 95°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
ICV	30mL	MP86729
CCV	30mL	MP86731
CRA	30mL	MP86733
Blank Spike	0.48mL	MP86722
Matrix Spike	0.48mL	MP86722

Chemical Used	ML/SAMPLE USED	Lot Number
HNO3/H2SO4(1:2)	2.25mL	MP85892
KMnO4 (5%)	4.5mL	MP85893
K2S2O8 (5%)	2.4mL	MP85894
Hydroxylamine HCL (12%)	1.8mL	MP85895
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
0.0 ppb	S0	30mL	MP86723
0.05 ppb	S0.05	30mL	MP86738
0.2 ppb	S0.2	30mL	MP86724
2.5 ppb	S2.5	30mL	MP86725
5.0 ppb	S5.0	30mL	MP86726
7.5 ppb	S7.5	30mL	MP86727
10.0 ppb	S10.0	30mL	MP86728
ICV	ICV	30mL	MP86729
ICB	ICB	30mL	MP86730
CCV	CCV	30mL	MP86731
CCB	CCB	30mL	MP86732
CRI	CRI	30mL	MP86733
CHK STD	CHK STD	30mL	MP86734

Extraction Conformance/Non-Conformance Comments:

N/A		
Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/12/25 16:00	<i>[Signature]</i>	<i>[Signature]</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Comment	Prep Pos
PB169219BL	PBW219	30	30	<2	N/A	1-1
PB169219BS	LCS219	30	30	<2	MP86722	2
Q2785-01MS	WATERMS	30	30	<2	MP86722	5
Q2785-01MSD	WATERMSD	30	30	<2	MP86722	6
Q2785-01DUP	WATERDUP	30	30	<2	N/A	4
Q2785-01	WATER	30	30	<2	N/A	3
Q2789-02	COMP	30	30	<2	N/A	7
Q2799-01	CC-071325-RW	30	30	<2	N/A	8

A
B
C
D
E
F
G
H
I
J