

CASE NARRATIVE

Earth Engineering Inc.

Project Name: 3313 BAY AVENUE

Project # N/A

Order ID # Q2802

Test Name: EPH_NF

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 08/07/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: EPH_NF. This data package contains results for EPH_NF.

C. Analytical Techniques:

The analysis were performed on instrument FID_C. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis were performed on instrument FID_E. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis of EPH_NFs was based on method NJEPH and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS {Q2798-01MS} with File ID: FC069626.D recoveries met the requirements for all compounds except for [n-Nonane (C9)- 37%] due to matrix interference.

The MS {Q2798-01MSD} with File ID: FC069627.D recoveries met the requirements for all compounds except for [n-Nonane (C9)- 39%] due to matrix interference.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____