

Continuing Calibration Report for SequenceID : FC080825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069620.D

Aliphatic C9-C12	7423392.000	60.000	3.298	6.596	123723.200	144679.273	14.485
Aliphatic C12-C16	4934649.000	40.000	6.597	10.000	123366.225	143435.206	13.992
Aliphatic C16-C21	6986048.000	60.000	10.001	13.369	116434.133	133561.231	12.823
Aliphatic C21-C28	8387531.000	80.000	13.370	17.034	104844.138	118307.822	11.380
Aliphatic C28-C40	10073631.000	120.000	17.035	22.011	83946.925	95315.213	11.927
Aliphatic EPH	37805251.000	360.000	3.298	22.011	105014.586	120373.027	12.759