

Continuing Calibration Report for SequenceID : FC080825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069631.D

Aliphatic C9-C12	7563783.000	60.000	3.298	6.596	126063.050	144679.273	12.867
Aliphatic C12-C16	4974222.000	40.000	6.597	10.000	124355.550	143435.206	13.302
Aliphatic C16-C21	7013845.000	60.000	10.001	13.369	116897.417	133561.231	12.477
Aliphatic C21-C28	8319419.000	80.000	13.370	17.034	103992.738	118307.822	12.100
Aliphatic C28-C40	10138718.000	120.000	17.035	22.011	84489.317	95315.213	11.358
Aliphatic EPH	38009987.000	360.000	3.298	22.011	105583.297	120373.027	12.287