

Continuing Calibration Report for SequenceID : FE080825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055211.D

Aliphatic C9-C12	7031222.000	60.000	3.325	6.960	117187.033	116782.597	-0.346
Aliphatic C12-C16	5061864.000	40.000	6.961	10.413	126546.600	125227.070	-1.054
Aliphatic C16-C21	7571571.000	60.000	10.414	13.792	126192.850	126404.887	0.168
Aliphatic C21-C28	10396476.000	80.000	13.793	17.464	129955.950	120505.719	-7.842
Aliphatic C28-C40	15499812.000	120.000	17.465	22.492	129165.100	113987.195	-13.315
Aliphatic EPH	45560945.000	360.000	3.325	22.492	126558.181	119220.147	-6.155