

Continuing Calibration Report for SequenceID : FE080825AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE055221.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 5963330.000  | 60.000  | 3.325  | 6.960  | 99388.833  | 116782.597 | 14.894 |
| Aliphatic C12-C16 | 4348051.000  | 40.000  | 6.961  | 10.413 | 108701.275 | 125227.070 | 13.197 |
| Aliphatic C16-C21 | 6703045.000  | 60.000  | 10.414 | 13.792 | 111717.417 | 126404.887 | 11.619 |
| Aliphatic C21-C28 | 9284570.000  | 80.000  | 13.793 | 17.464 | 116057.125 | 120505.719 | 3.692  |
| Aliphatic C28-C40 | 13756320.000 | 120.000 | 17.465 | 22.492 | 114636.000 | 113987.195 | -0.569 |
| Aliphatic EPH     | 40055316.000 | 360.000 | 3.325  | 22.492 | 111264.767 | 119220.147 | 6.673  |