

## Report of Analysis

Client:	VERINA CONSULTING GROUP, LLC	Date Collected:	08/07/25 10:48
Project:	Rotor Clip NJ WTD - 2025	Date Received:	08/07/25
Client Sample ID:	WATER-TREATMENT DISCHARGE	SDG No.:	Q2804
Lab Sample ID:	Q2804-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	0.11		1	0.030	0.10	mg/L	08/11/25 15:00	08/12/25 10:19	SM 4500-NH3 B plus G-21
Residual Chlorine	0.023	HU	1	0.023	0.10	mg/L		08/07/25 16:25	SM 4500-Cl G-11

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits