

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2804
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136744

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136744BS							
Residual Chlorine	mg/L	0.4	0.40		100	1	90-110	08/07/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2804
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136784

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169199BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	08/12/2025