

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2804

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	793	800	99	90 - 110	P	08/11/2025	16:44	LB136779
	Copper	989	1000	99	90 - 110	P	08/11/2025	16:44	LB136779
	Nickel	1940	2000	97	90 - 110	P	08/11/2025	16:44	LB136779
	Zinc	1920	2000	96	90 - 110	P	08/11/2025	16:44	LB136779

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2804

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	9.85	10.0	98	80 - 120	P	08/11/2025	16:49	LB136779
	Copper	21.3	20.0	107	80 - 120	P	08/11/2025	16:49	LB136779
	Nickel	39.9	40.0	100	80 - 120	P	08/11/2025	16:49	LB136779
	Zinc	41.2	40.0	103	80 - 120	P	08/11/2025	16:49	LB136779

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2804

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	972	1000	97	90 - 110	P	08/11/2025	17:19	LB136779
	Copper	1240	1250	99	90 - 110	P	08/11/2025	17:19	LB136779
	Nickel	2440	2500	98	90 - 110	P	08/11/2025	17:19	LB136779
	Zinc	2440	2500	98	90 - 110	P	08/11/2025	17:19	LB136779
CCV02	Chromium	992	1000	99	90 - 110	P	08/11/2025	17:44	LB136779
	Copper	1250	1250	100	90 - 110	P	08/11/2025	17:44	LB136779
	Nickel	2470	2500	99	90 - 110	P	08/11/2025	17:44	LB136779
	Zinc	2450	2500	98	90 - 110	P	08/11/2025	17:44	LB136779
CCV03	Chromium	977	1000	98	90 - 110	P	08/11/2025	18:26	LB136779
	Copper	1240	1250	99	90 - 110	P	08/11/2025	18:26	LB136779
	Nickel	2450	2500	98	90 - 110	P	08/11/2025	18:26	LB136779
	Zinc	2450	2500	98	90 - 110	P	08/11/2025	18:26	LB136779
CCV04	Chromium	979	1000	98	90 - 110	P	08/11/2025	18:59	LB136779
	Copper	1230	1250	98	90 - 110	P	08/11/2025	18:59	LB136779
	Nickel	2430	2500	97	90 - 110	P	08/11/2025	18:59	LB136779
	Zinc	2440	2500	98	90 - 110	P	08/11/2025	18:59	LB136779