



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2804

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	9.63	10.0	96	65 - 135	P	08/11/2025	16:58	LB136779
	Copper	20.7	20.0	103	65 - 135	P	08/11/2025	16:58	LB136779
	Nickel	39.0	40.0	98	65 - 135	P	08/11/2025	16:58	LB136779
	Zinc	39.4	40.0	99	65 - 135	P	08/11/2025	16:58	LB136779