

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2804

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169154BL	WATER			Batch Number:	PB169154		Prep Date:	08/08/2025	
	Chromium	1.06	<2.5	U	5.00	P	08/11/2025	17:52	LB136779
	Copper	2.30	<5	U	10.0	P	08/11/2025	17:52	LB136779
	Nickel	1.53	<10	U	20.0	P	08/11/2025	17:52	LB136779
	Zinc	8.33	<10	U	20.0	P	08/11/2025	17:52	LB136779