

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2804

Contract: VERI01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	56.8	52.0	109	42	62	08/11/2025	17:02	LB136779
	Copper	-0.68	2.0	34	-18	22	08/11/2025	17:02	LB136779
	Nickel	5.25	2.0	262	-38	42	08/11/2025	17:02	LB136779
	Zinc	2.14			-40	40	08/11/2025	17:02	LB136779
ICSAB01	Chromium	534	542	98	460	624	08/11/2025	17:06	LB136779
	Copper	463	511	91	434	588	08/11/2025	17:06	LB136779
	Nickel	949	954	100	810	1100	08/11/2025	17:06	LB136779
	Zinc	987	952	104	809	1095	08/11/2025	17:06	LB136779