

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q2804
contract: VERI01 **lab code:** ACE
matrix: Water **sample id:** Q2805-02 **client id:** RW8-SP303-20250807MS
Percent Solids for Sample: NA **Spiked ID:** Q2805-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	90 - 113	206		5.00	U	200	103		P
Copper	ug/L	86 - 114	162		6.11	J	150	104		P
Nickel	ug/L	88 - 113	254		3.69	J	250	100		P
Zinc	ug/L	87 - 115	132		29.6		100	103		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>VERINA CONSULTING GROUP, LLC</u>	level: <u>low</u>	sdg no.: <u>Q2804</u>
contract: <u>VERI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2805-02</u>	client id: <u>RW8-SP303-20250807MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2805-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	90 - 113	203		5.00	U	200	102		P
Copper	ug/L	86 - 114	154		6.11	J	150	99		P
Nickel	ug/L	88 - 113	241		3.69	J	250	95		P
Zinc	ug/L	87 - 115	130		29.6		100	101		P