

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q2804
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2805-02 **Client ID:** RW8-SP303-20250807DUP
Percent Solids for Sample: NA **Duplicate ID** Q2805-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	5.00	U	5.00	U			P
Copper	ug/L	20	6.11	J	6.10	J	0		P
Nickel	ug/L	20	3.69	J	3.86	J	5		P
Zinc	ug/L	20	29.6		30.2		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q2804
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2805-02MS **Client ID:** RW8-SP303-20250807MSD
Percent Solids for Sample: NA **Duplicate ID** Q2805-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	206		203		1		P
Copper	ug/L	20	162		154		5		P
Nickel	ug/L	20	254		241		5		P
Zinc	ug/L	20	132		130		2		P