

## **DATA PACKAGE GENERAL CHEMISTRY**

**PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13**

**TETRA TECH NUS, INC.**

**661 Andersen Drive**

**Suite 200**

**Pittsburgh, PA - 15220-2745**

**Phone No: 412-921-7090**

**ORDER ID : Q2805**

**ATTENTION : Ernie Wu**



**Laboratory Certification ID # 20012**



1) GENERAL CHEMISTRY DATA	2
2) Signature Page	3
3) Case Narrative	4
4) Qualifier Page	5
5) Conformance/Non Conformance	6
6) QA Checklist	7
7) Chronicle	8
8) Sample Data	9
8.1) RW8-SP303-20250807	10
9) QC Data Summary For Genchem	11
9.1) Preparation Blank Summary	12
9.2) Duplicate Sample Summary	13
9.3) Laboratory Control Sample Summary	15
10) GENCHEM RAW DATA	17
10.1) GENCHEM RAW DATA - ANALYTICAL	18
10.1.1) LB136751	18
10.1.2) LB136765	20
11) Analytical Runlogs	22
12) Standard Prep Logs	24
13) Shipping Document	26
13.1) Chain Of Custody	27
13.2) Lab Certificate	28

1
2
3
4
5
6
7
8
9
10
11
12
13

## Cover Page

**Order ID :** Q2805

**Project ID :** NWIRP Bethpage 112G08005-WE13

**Client :** Tetra Tech NUS, Inc.

**Lab Sample Number**

Q2805-01  
Q2805-02

**Client Sample Number**

RW8-SP100-20250807  
RW8-SP303-20250807

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 12:41 pm, Aug 19, 2025*

Date: 8/13/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Tetra Tech NUS, Inc.**

**Project Name: NWIRP Bethpage 112G08005-WE13**

**Project Manager : Ernie Wu**

**Order ID # Q2805**

**Test Name: TDS,TSS**

### **A. Number of Samples and Date of Receipt:**

1 Water sample was received on 08/08/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: TDS,TSS. This data package contains results for TDS,TSS.

### **C. Analytical Techniques:**

The analysis of TDS was based on method SM2540 C and The analysis of TSS was based on method SM2540 D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

**APPROVED**

Signature\_\_\_\_\_

*By Nimisha Pandya, QA/QC Supervisor at 12:41 pm, Aug 19, 2025*

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

<b>J</b>	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
<b>U</b>	Indicates the analyte was analyzed for, but not detected.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>E</b>	Indicates the reported value is estimated because of the presence of interference
<b>M</b>	Indicates Duplicate injection precision not met.
<b>N</b>	Indicates the spiked sample recovery is not within control limits.
<b>S</b>	Indicates the reported value was determined by the Method of Standard Addition (MSA).
<b>*</b>	Indicates that the duplicate analysis is not within control limits.
<b>+</b>	Indicates the correlation coefficient for the MSA is less than 0.995.
<b>D</b>	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
<b>M</b>	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
<b>OR</b>	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements
<b>H</b>	Sample Analysis Out Of Hold Time

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

ORDER ID: Q2805

MATRIX: Water

METHOD: SM2540 C,SM2540 D

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. The Blank Spike met requirements for all compounds.			✓
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

QA REVIEW

REVIEWED

Sohil Jodhani, QA/QC Director , 8/19/2025, 10:30:15 AM

APPENDIX A

**QA REVIEW GENERAL DOCUMENTATION**

Project #: Q2805

Completed

For thorough review, the report must have the following:

**GENERAL:**

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

**COVER PAGE:**

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

**CHAIN OF CUSTODY:**

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

**ANALYTICAL:**

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 08/13/2025

### LAB CHRONICLE

<b>OrderID:</b>	Q2805	<b>OrderDate:</b>	8/8/2025 9:48:00 AM
<b>Client:</b>	Tetra Tech NUS, Inc.	<b>Project:</b>	NWIRP Bethpage 112G08005-WE13
<b>Contact:</b>	Ernie Wu	<b>Location:</b>	J22

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2805-02	RW8-SP303-2025080 7	WATER			08/07/25 13:13			08/08/25
			TDS	SM2540 C			08/08/25 17:00	
			TSS	SM2540 D			08/11/25 10:00	





# SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13

## Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	08/07/25 13:13
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	08/08/25
Client Sample ID:	RW8-SP303-20250807	SDG No.:	Q2805
Lab Sample ID:	Q2805-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	3.00	J	1	1.00	10.0	10.0	mg/L		08/08/25 17:00	SM 2540 C-20
TSS	4.00	U	1	1.00	4.00	4.00	mg/L		08/11/25 10:00	SM 2540 D-20

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

# QC RESULT SUMMARY

## Preparation Blank Summary

**Client:** Tetra Tech NUS, Inc.

**SDG No.:** Q2805

**Project:** NWIRP Bethpage 112G08005-WE13

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: <b>LB136751BL</b>							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	08/08/2025
Sample ID: <b>LB136765BL</b>							
TSS	mg/L	< 2.0000	2.0000	U	1	4	08/11/2025

### Duplicate Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2805
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Sample ID:</b>	Q2789-02
<b>Client ID:</b>	CompDUP	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	377		387		1	2.62		08/11/2025

### Duplicate Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2805
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Sample ID:</b>	Q2805-02
<b>Client ID:</b>	RW8-SP303-20250807DUP	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TDS	mg/L	+/-5	3.00	J	3.00	J	1	0		08/08/2025

### Laboratory Control Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2805
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Run No.:</b>	LB136751

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136751BS							
TDS	mg/L	100	95.0		95	1	90-110	08/08/2025

### Laboratory Control Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2805
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Run No.:</b>	LB136765

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136765BS							
TSS	mg/L	550	531		96	1	90-110	08/11/2025





# RAW DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

# TOTAL Dissolved Solids - SM2540C

Run Number: LB136751

Date: 08/11/2025

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC-SC-6

Filter ID: 17416528

EMPTY DISH				EMPTY DISH				OvenID:	WC OVEN-1	Thermo ID:	WET OVEN#1
TEMP IN:	104 °C	08/08/2025	11:00	TEMP OUT:	104 °C	08/08/2025	12:00				
TEMP1 IN:	104 °C	08/08/2025	12:30	TEMP1 OUT:	104 °C	08/08/2025	13:30	OvenID1:	WC OVEN-1	Thermo ID1:	WET OVEN#1
TEMP2 IN:	103 °C	08/08/2025	17:00	TEMP2 OUT:	103 °C	08/11/2025	07:00	OvenID2:	WC OVEN-2	Thermo ID2:	WET OVEN#2
TEMP3 IN:	180 °C	08/11/2025	07:05	TEMP3 OUT:	180 °C	08/11/2025	09:00	OvenID3:	WC OVEN-2	Thermo ID3:	WET OVEN#2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180(±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180(±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180(±2) °C (g)	Weight (g)	Result mg/L
1	LB136751BL	LB136751BL	145.6983	145.6983	100	145.6983	145.6983	145.6983	0.0000	0
2	LB136751BS	LB136751BS	104.8563	104.8563	100	104.8658	104.8658	104.8658	0.0095	95
3	Q2805-02	RW8-SP303-20250807	110.2296	110.2296	100	110.2299	110.2299	110.2299	0.0003	3
4	Q2805-02DUP	RW8-SP303-20250807DUP	103.6217	103.6217	100	103.6220	103.6220	103.6220	0.0003	3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180(±2) °C(g)

D = Weight (g)

Q2805-GENCHEM

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

# WORKLIST(Hardcopy Internal Chain)

VB136f51

WorkList Name : tds q2805

WorkList ID : 191180

Department : Wet-Chemistry

Date : 08-08-2025 12:49:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2805-02	B	RW8-SP303-20250807	Water	TDS	Cool 4 deg C	TETR06	J22	08/07/2025 SM2540 C

Date/Time 08/08/25 13:110  
 Raw Sample Received by: sb wdy  
 Raw Sample Relinquished by: JT (sm)

Date/Time 08/08/25 17:30  
 Raw Sample Received by: JT (sm)  
 Raw Sample Relinquished by: sb wdy

# TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: rubina

ANALYST: jignesh

Date: 08/08/2025

Run Number: LB136765

BalanceID: WC-SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 08/08/2025 14:00 TEMP1 OUT: 103 °C 08/08/2025 15:30  
 TEMP2 IN: 104 °C 08/08/2025 16:00 TEMP2 OUT: 103 °C 08/08/2025 17:00  
 TEMP3 IN: 104 °C 08/11/2025 10:00 TEMP3 OUT: 103 °C 08/11/2025 11:30  
 TEMP4 IN: 104 °C 08/11/2025 12:00 TEMP4 OUT: 103 °C 08/11/2025 13:35

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB136765BL	LB136765BL	1.5206	1.5207	100	1.5207	1.5207	1.5207	0.0000	0
2	LB136765BS	LB136765BS	1.4853	1.4854	100	1.5385	1.5385	1.5385	0.0531	531
3	Q2789-02	Comp	1.4883	1.4883	70	1.5147	1.5147	1.5147	0.0264	377.1
4	Q2789-02DUP	CompDUP	1.4859	1.4859	70	1.5129	1.5130	1.5130	0.0271	387.1
5	Q2805-02	RW8-SP303-20250807	1.4726	1.4727	1800	1.4730	1.4731	1.4731	0.0004	0.2
6	Q2810-01	MH-892025	1.4753	1.4753	300	1.5575	1.5575	1.5575	0.0822	274
7	Q2811-02	EFF-WASTE WATER	1.4686	1.4686	750	1.4911	1.4911	1.4911	0.0225	30
8	Q2813-01	EFFLUENT	1.4912	1.4912	20	1.6767	1.6767	1.6767	0.1855	9275
9	Q2813-04	AERATION	1.5005	1.5005	20	1.5670	1.5670	1.5670	0.0665	3325

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L =  $\frac{D}{A} \times 1000 \times 1000$

## WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS Q2810      WorkList ID : 191188      Department : Wet-Chemistry      Date : 08-11-2025 07:52:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2789-02	B Comp	Water	TSS	Cool 4 deg C	ARAM01	J21	08/06/2025	SM2540 D
Q2805-02	BL RW8-SP303-20250807	Water	TSS	Cool 4 deg C	TETR06	J22	08/07/2025	SM2540 D
Q2810-01	D MH-892025	Water	TSS	Cool 4 deg C	EURO03	D21	08/08/2025	SM2540 D
Q2811-02	B EFF-WASTE WATER	Water	TSS	Cool 4 deg C	ARDM01	J21	08/08/2025	SM2540 D
Q2813-01	D EFFLUENT	Water	TSS	Cool 4 deg C	HOLL01	J41	08/08/2025	SM2540 D
Q2813-04	AERATION	Water	TSS	Cool 4 deg C	HOLL01	J41	08/08/2025	SM2540 D

Date/Time 08-11-25 08:00  
 Raw Sample Received by: SB wcl  
 Raw Sample Relinquished by: JDCSM

Date/Time 08-11-25 14:30  
 Raw Sample Received by: JDCSM  
 Raw Sample Relinquished by: SB wcl

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136751**

Review By	jignesh	Review On	8/11/2025 11:53:29 AM
Supervise By	Iwona	Supervise On	8/11/2025 12:52:45 PM
SubDirectory	LB136751	Test	TDS
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136751BL	LB136751BL	MB	08/08/25 17:00		jignesh	OK
2	LB136751BS	LB136751BS	LCS	08/08/25 17:00		jignesh	OK
3	Q2805-02	RW8-SP303-2025080	SAM	08/08/25 17:00		jignesh	OK
4	Q2805-02DUP	RW8-SP303-2025080	DUP	08/08/25 17:00		jignesh	OK

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136765**

Review By	jignesh	Review On	8/12/2025 2:27:51 PM
Supervise By	rubina	Supervise On	8/12/2025 2:30:28 PM
SubDirectory	LB136765	Test	TSS
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136765BL	LB136765BL	MB	08/11/25 10:00		jignesh	OK
2	LB136765BS	LB136765BS	LCS	08/11/25 10:00		jignesh	OK
3	Q2789-02	Comp	SAM	08/11/25 10:00		jignesh	OK
4	Q2789-02DUP	CompDUP	DUP	08/11/25 10:00		jignesh	OK
5	Q2805-02	RW8-SP303-2025080	SAM	08/11/25 10:00		jignesh	OK
6	Q2810-01	MH-892025	SAM	08/11/25 10:00		jignesh	OK
7	Q2811-02	EFF-WASTE WATER	SAM	08/11/25 10:00		jignesh	OK
8	Q2813-01	EFFLUENT	SAM	08/11/25 10:00		jignesh	OK
9	Q2813-04	AERATION	SAM	08/11/25 10:00		jignesh	OK

### Prep Standard - Chemical Standard Summary

**Order ID :** Q2805

**Test :** TDS,TSS

**Prepbatch ID :**

**Sequence ID/Qc Batch ID:** LB136751, LB136765,

**Standard ID :**

**Chemical ID :**

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13



### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
				/	/	

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13



# SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

<b>CHEMTECH</b> CHAIN OF CUSTODY RECORD		284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 78-8922 www.chemtech.net		Chemtech Project Number: <span style="font-size: 1.5em; margin-left: 10px;">Q 2805</span>																																											
				COC Number:																																											
<b>CLIENT INFORMATION</b>				<b>PROJECT INFORMATION</b>				<b>BILLING INFORMATION</b>																																							
COMPANY: Tetra Tech				PROJECT NAME: NWIRP Bethpage				BILL TO:					PO#																																		
ADDRESS: 4433 Corporation Ln, Suite 300				PROJECT #: 112G08005-WE13      LOCATION: RW8				ADDRESS:																																							
CITY: Virginia Beach      STATE: VA      ZIP: 23462				PROJECT MANAGER: Ernie Wu				CITY:					STATE:      ZIP:																																		
ATTENTION: Ernie Wu				E-MAIL: ernie.wu@tetratech.com				ATTENTION:					PHONE:																																		
PHONE: 757-466-4901      FAX: 757-461-4148				PHONE: 757-466-4901      FAX: 757-461-4148																																											
<b>DATA TURNAROUND INFORMATION</b>				<b>DATA DELIVERABLE INFORMATION</b>				<b>ANALYSIS</b>																																							
FAX: _____ 10 _____ DAYS* HARD COPY: _____ 10 _____ DAYS* EDD _____ 10 _____ DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS				<input type="checkbox"/> RESEULTS ONLY <input type="checkbox"/> USEPA CLP <input type="checkbox"/> RESULTS + QC <input type="checkbox"/> New York State ASP "B" <input type="checkbox"/> New Jersey REDUCED <input type="checkbox"/> New York State ASP "A" <input type="checkbox"/> New Jersey CLP <input type="checkbox"/> Other _____ <input type="checkbox"/> EDD Format _____				<table border="1" style="width:100%; border-collapse: collapse; text-align: center;"> <tr> <td style="writing-mode: vertical-rl; transform: rotate(180deg); font-size: 0.8em;">1,4-Dioxane SW846 8270 SIM</td> <td style="writing-mode: vertical-rl; transform: rotate(180deg); font-size: 0.8em;">Iron, Total</td> <td style="writing-mode: vertical-rl; transform: rotate(180deg); font-size: 0.8em;">TSS</td> <td style="writing-mode: vertical-rl; transform: rotate(180deg); font-size: 0.8em;">TDS</td> <td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td> </tr> <tr> <td>1</td><td>2</td><td>3</td><td>4</td><td>5</td><td>6</td><td>7</td><td>8</td><td>9</td><td></td><td></td><td></td><td></td><td></td><td></td> </tr> </table>										1,4-Dioxane SW846 8270 SIM	Iron, Total	TSS	TDS												1	2	3	4	5	6	7	8	9						
																		1,4-Dioxane SW846 8270 SIM	Iron, Total	TSS	TDS																										
1	2	3	4	5	6	7	8	9																																							
<b>PRESERVATIVES</b>										<b>COMMENTS</b>																																					
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION			SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										<-- Specify Preservatives A-HCl      B-HNO3 C-H2SO4      D-NaOH E-ICE      F-Other																												
				COMP	GRAB	DATE	TIME	1		2	3	4	5	6	7	8	9																														
1.	RW8-SP100-20250807			GW		x	8/7/25	13:05	2	x	x							pH 1.3																													
2.	RW8-SP303-20250807			GW		x	8/7/25	13:13	4	x	x	x	x					pH 1.3																													
3.																																															
4.																																															
5.																																															
6.																																															
7.																																															
8.																																															
9.																																															
10.																																															
<b>SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY</b>																																															
RELINQUISHED BY SAMPLER		DATE/TIME		RECEIVED BY				Conditions of bottles or coolers at receipt: <input type="checkbox"/> Compliant <input type="checkbox"/> Non Compliant <input type="checkbox"/> Cooler Temp <u>2.1°C</u> MeOH extraction requires an additional 4oz. Jar for percent solid <input type="checkbox"/> Ice in Cooler? <u>yes</u> Comments: <span style="float: right; margin-top: 10px;">IR Gun # 1</span>																																							
1.		8/7/25/1504		1.																																											
RELINQUISHED BY		DATE/TIME		RECEIVED BY																																											
2.		8/8/25		2.				SHIPPED VIA: CLIENT: <input type="checkbox"/> Hand Delivered <input type="checkbox"/> Overnight CHEMTECH: <input type="checkbox"/> Picked Up <input type="checkbox"/> Overnight																																							
RELINQUISHED BY		DATE/TIME		RECEIVED FOR LAB BY																																											
3.				3.				<b>Shipment Complete</b> <input type="checkbox"/> YES <input type="checkbox"/> NO																																							
Page _____ of _____																																															
WHITE - CHEMTECH COPY FOR RETURN TO CLIENT    YELLOW - CHEMTECH COPY    PINK - SAMPLER COPY																																															

### Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488