

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2805

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3930	4000	98	90 - 110	P	08/11/2025	16:44	LB136779

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2805

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	84.3	100	84	80 - 120	P	08/11/2025	16:49	LB136779

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2805

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4850	5000	97	90 - 110	P	08/11/2025	17:19	LB136779
CCV02	Iron	4960	5000	99	90 - 110	P	08/11/2025	17:44	LB136779
CCV03	Iron	4910	5000	98	90 - 110	P	08/11/2025	18:26	LB136779
CCV04	Iron	4970	5000	99	90 - 110	P	08/11/2025	18:59	LB136779