

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tetra Tech NUS, Inc. level: low sdg no.: Q2805
contract: TETR06 lab code: ACE
matrix: Water sample id: Q2805-02 client id: RW8-SP303-20250807MS
Percent Solids for Sample: NA Spiked ID: Q2805-02MS Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1540		33.3	J	1500	100		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q2805
contract: TETR06 **lab code:** ACE
matrix: Water **sample id:** Q2805-02 **client id:** RW8-SP303-20250807MSD
Percent Solids for Sample: NA **Spiked ID:** Q2805-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1510		33.3	J	1500	98		P