

LAB CHRONICLE

OrderID:	Q2805	OrderDate:	8/8/2025 9:48:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	J22

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2805-01	RW8-SP100-20250807	Water			08/07/25			08/08/25
			SVOC-SIMGroup1	8270-Modified		08/12/25	08/13/25	
Q2805-02	RW8-SP303-20250807	Water			08/07/25			08/08/25
			SVOC-SIMGroup1	8270-Modified		08/12/25	08/13/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2805
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				0.000					
Total Svoc :					0.00				
Total Concentration:					0.00				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	08/07/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	08/08/25
Client Sample ID:	RW8-SP100-20250807	SDG No.:	Q2805
Lab Sample ID:	Q2805-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037590.D	1	08/12/25 08:52	08/13/25 11:42	PB169222

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	UQ	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		96%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		93%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.41		53 - 106		102%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.46		58 - 132		116%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2070	7.717				
1146-65-2	Naphthalene-d8	4940	10.498				
15067-26-2	Acenaphthene-d10	2390	14.345				
1517-22-2	Phenanthrene-d10	4910	17.087				
1719-03-5	Chrysene-d12	4350	21.268				
1520-96-3	Perylene-d12	3800	23.508				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	08/07/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	08/08/25
Client Sample ID:	RW8-SP303-20250807	SDG No.:	Q2805
Lab Sample ID:	Q2805-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037591.D	1	08/12/25 08:52	08/13/25 12:18	PB169222

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	UQ	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.36		30 - 150		89%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		92%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		112%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2150	7.717				
1146-65-2	Naphthalene-d8	5100	10.498				
15067-26-2	Acenaphthene-d10	2480	14.345				
1517-22-2	Phenanthrene-d10	4960	17.086				
1719-03-5	Chrysene-d12	4080	21.277				
1520-96-3	Perylene-d12	3500	23.507				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: **Q2805**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **8270-Modified**

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169222BL	PB169222BL	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.32	81		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.35	86		53	106
		Terphenyl-d14	0.4	0.38	94		58	132
PB169222BS	PB169222BS	2-Methylnaphthalene-d10	0.4	0.34	84		30	150
		Fluoranthene-d10	0.4	0.30	75		30	150
		Nitrobenzene-d5	0.4	0.34	84		55	111
		2-Fluorobiphenyl	0.4	0.36	89		53	106
		Terphenyl-d14	0.4	0.37	93		58	132
PB169222BSD	PB169222BSD	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.32	80		30	150
		Nitrobenzene-d5	0.4	0.36	91		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.41	101		58	132
Q2805-01	RW8-SP100-20250807	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.39	96		30	150
		Nitrobenzene-d5	0.4	0.37	93		55	111
		2-Fluorobiphenyl	0.4	0.41	102		53	106
		Terphenyl-d14	0.4	0.46	116		58	132
Q2805-02	RW8-SP303-20250807	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.36	89		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.37	92		53	106
		Terphenyl-d14	0.4	0.45	112		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2805

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037596.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169222BS	1,4-Dioxane	0.4	0.27	ug/L	68		*		70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2805

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037597.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169222BSD	1,4-Dioxane	0.4	0.29	ug/L	73	7			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB169222BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2805

Lab File ID: BN037589.D

Lab Sample ID: PB169222BL

Instrument ID: BNA_N

Date Extracted: 08/12/2025

Matrix: (soil/water) Water

Date Analyzed: 08/13/2025

Level: (low/med) LOW

Time Analyzed: 11:05

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB169222BS	PB169222BS	BN037596.D	08/13/2025
RW8-SP100-20250807	Q2805-01	BN037590.D	08/13/2025
RW8-SP303-20250807	Q2805-02	BN037591.D	08/13/2025
PB169222BSD	PB169222BSD	BN037597.D	08/13/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037576.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2805
DFTPP Injection Date: 08/12/2025
DFTPP Injection Time: 15:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	87.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.2 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037578.D	08/12/2025	16:26
SSTDICC0.2	SSTDICC0.2	BN037579.D	08/12/2025	17:03
SSTDICCC0.4	SSTDICCC0.4	BN037580.D	08/12/2025	17:39
SSTDICC0.8	SSTDICC0.8	BN037581.D	08/12/2025	18:16
SSTDICC1.6	SSTDICC1.6	BN037582.D	08/12/2025	18:52
SSTDICC3.2	SSTDICC3.2	BN037583.D	08/12/2025	19:29
SSTDICC5.0	SSTDICC5.0	BN037584.D	08/12/2025	20:05

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037587.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2805
DFTPP Injection Date: 08/13/2025
DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	74.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (21) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037588.D	08/13/2025	10:28
PB169222BL	PB169222BL	BN037589.D	08/13/2025	11:05
RW8-SP100-20250807	Q2805-01	BN037590.D	08/13/2025	11:42
RW8-SP303-20250807	Q2805-02	BN037591.D	08/13/2025	12:18
PB169222BS	PB169222BS	BN037596.D	08/13/2025	15:22
PB169222BSD	PB169222BSD	BN037597.D	08/13/2025	15:58
SSTDCCC0.4EC	SSTDCCC0.4	BN037598.D	08/13/2025	16:35

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2805

Client ID : SSTDCCC0.4

Date Analyzed: 08/13/2025

Lab File ID: BN037588.D

Time Analyzed: 10:28

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2728	7.717	6843	10.50	3392	14.35
UPPER LIMIT	5456	8.217	13686	10.998	6784	14.845
LOWER LIMIT	1364	7.217	3421.5	9.998	1696	13.845
EPA SAMPLE NO.						
01 PB169222BL	2235	7.72	5344	10.50	2468	14.35
02 RW8-SP100-20250807	2074	7.72	4942	10.50	2387	14.35
03 RW8-SP303-20250807	2147	7.72	5102	10.50	2477	14.35
04 PB169222BS	3068	7.72	7591	10.50	3534	14.35
05 PB169222BSD	2742	7.72	6598	10.50	3080	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2805

Client ID: SSTDCCC0.4

Date Analyzed: 08/13/2025

Lab File ID: BN037588.D

Time Analyzed: 10:28

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6530	17.086	6467	21.268	5771	23.505
UPPER LIMIT	13060	17.586	12934	21.768	11542	24.005
LOWER LIMIT	3265	16.586	3233.5	20.768	2885.5	23.005
EPA SAMPLE NO.						
01 PB169222BL	4726	17.10	3815	21.27	3770	23.51
02 RW8-SP100-20250807	4913	17.09	4346	21.27	3800	23.51
03 RW8-SP303-20250807	4957	17.09	4082	21.28	3495	23.51
04 PB169222BS	6432	17.09	4873	21.28	4192	23.51
05 PB169222BSD	5659	17.09	4216	21.28	3616	23.51

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169222BL		SDG No.:	Q2805
Lab Sample ID:	PB169222BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037589.D	1	08/12/25 08:52	08/13/25 11:05	PB169222

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.32		30 - 150		81%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		86%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		94%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2240	7.717				
1146-65-2	Naphthalene-d8	5340	10.498				
15067-26-2	Acenaphthene-d10	2470	14.345				
1517-22-2	Phenanthrene-d10	4730	17.099				
1719-03-5	Chrysene-d12	3820	21.268				
1520-96-3	Perylene-d12	3770	23.508				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB169222BS		SDG No.:	Q2805	
Lab Sample ID:	PB169222BS		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037596.D	1	08/12/25 08:52	08/13/25 15:22	PB169222

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.27		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.34		30 - 150		84%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 - 150		75%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		84%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		89%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.37		58 - 132		93%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	3070	7.717				
1146-65-2	Naphthalene-d8	7590	10.498				
15067-26-2	Acenaphthene-d10	3530	14.345				
1517-22-2	Phenanthrene-d10	6430	17.086				
1719-03-5	Chrysene-d12	4870	21.277				
1520-96-3	Perylene-d12	4190	23.51				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169222BSD		SDG No.:	Q2805
Lab Sample ID:	PB169222BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037597.D	1	08/12/25 08:52	08/13/25 15:58	PB169222

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.29		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.32		30 - 150		80%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		55 - 111		91%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		97%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.41		58 - 132		101%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2740	7.717				
1146-65-2	Naphthalene-d8	6600	10.498				
15067-26-2	Acenaphthene-d10	3080	14.345				
1517-22-2	Phenanthrene-d10	5660	17.086				
1719-03-5	Chrysene-d12	4220	21.277				
1520-96-3	Perylene-d12	3620	23.505				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN081225.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 13 05:00:58 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037578.D 0.2 =BN037579.D 0.4 =BN037580.D 0.8 =BN037581.D 1.6 =BN037582.D 3.2 =BN037583.D 5 =BN037584.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.413	0.395	0.368	0.397	0.375	0.348	0.383		6.08
3)	n-Nitrosodimet...	0.471	0.488	0.471	0.493	0.503	0.508	0.489		3.19
4) S	2-Fluorophenol	0.922	0.945	0.888	0.823	0.886	0.906	0.977	0.907	5.41
5) S	Phenol-d6	1.045	1.051	1.044	0.983	1.198	1.117	1.201	1.091	7.66
6)	bis(2-Chloroet...	0.935	0.987	0.976	0.936	1.020	1.008	1.022	0.983	3.75
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.270	0.257	0.262	0.264	0.292	0.301	0.326	0.282	9.03
9)	Naphthalene	1.044	1.050	1.037	1.012	1.089	1.096	1.127	1.065	3.78
10)	Hexachlorobuta...	0.252	0.258	0.262	0.253	0.267	0.264	0.265	0.260	2.22
11) SURR2	Methylnaphth...	0.488	0.492	0.508	0.493	0.543	0.578	0.705	0.544	14.39
12)	2-Methylnaphth...	0.589	0.631	0.635	0.629	0.701	0.731	0.760	0.668	9.39
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.139	0.149	0.156	0.160	0.185	0.203	0.234	0.175	19.39
15) S	2-Fluorobiphenyl	2.226	2.243	2.300	2.251	2.341	2.391	2.440	2.313	3.50
16)	Acenaphthylene	1.740	1.760	1.710	1.653	1.850	1.858	1.980	1.793	6.15
17)	Acenaphthene	1.144	1.150	1.172	1.154	1.283	1.286	1.348	1.220	6.86
18)	Fluorene	1.466	1.510	1.524	1.521	1.686	1.693	1.770	1.596	7.38
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.042	0.046	0.047	0.061	0.070		0.053		22.37
21)	4-Bromophenyl-...	0.226	0.236	0.234	0.237	0.265	0.271	0.292	0.251	9.77
22)	Hexachlorobenzene	0.360	0.360	0.356	0.332	0.364	0.354	0.370	0.356	3.37
23)	Atrazine	0.127	0.126	0.130	0.134	0.160	0.176		0.142	14.76
24)	Pentachlorophenol		0.108	0.113	0.113	0.139	0.152		0.125	15.36
25)	Phenanthrene	1.146	1.162	1.170	1.138	1.281	1.269	1.347	1.216	6.72
26)	Anthracene	0.950	0.985	0.995	0.988	1.145	1.186	1.286	1.077	11.93
27) SURR	Fluoranthene-d10	0.932	0.966	0.954	0.935	1.050	1.110	1.419	1.052	16.61
28)	Fluoranthene	1.241	1.264	1.299	1.291	1.503	1.533	1.648	1.397	11.53
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.539	1.486	1.464	1.403	1.502	1.520	1.552	1.495	3.38
31) S	Terphenyl-d14	0.814	0.801	0.796	0.763	0.833	0.853	0.901	0.823	5.45
32)	Benzo(a)anthra...	1.266	1.275	1.263	1.253	1.351	1.458	1.487	1.336	7.40
33)	Chrysene	1.547	1.515	1.433	1.379	1.511	1.489	1.551	1.489	4.21
34)	Bis(2-ethylhex...		0.518	0.522	0.483	0.522	0.593	0.671	0.551	12.47
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN081225.M

36)	Indeno(1,2,3-c...	1.446	1.462	1.612	1.568	1.816	1.883	2.011	1.686	13.01
37)	Benzo(b)fluora...	1.351	1.375	1.338	1.432	1.615	1.674	1.825	1.516	12.52
38)	Benzo(k)fluora...	1.597	1.580	1.651	1.628	1.821	1.797	1.898	1.710	7.36
39) C	Benzo(a)pyrene	1.146	1.136	1.149	1.157	1.322	1.382	1.508	1.257	11.77
40)	Dibenzo(a,h)an...	1.023	1.118	1.222	1.215	1.439	1.516	1.621	1.308	16.85
41)	Benzo(g,h,i)pe...	1.207	1.274	1.262	1.261	1.467	1.532	1.643	1.378	12.17

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q2805
Instrument ID: BNA_N Calibration Date/Time: 08/13/2025 10:28
Lab File ID: BN037588.D Init. Calib. Date(s): 08/12/2025 08/12/2025
EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 16:26 20:05
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.508		-6.6	20.0
Fluoranthene-d10	1.052	0.929		-11.7	20.0
2-Fluorophenol	0.907	0.923		1.8	20.0
Phenol-d6	1.091	1.135		4.0	20.0
Nitrobenzene-d5	0.282	0.265		-6.0	20.0
2-Fluorobiphenyl	2.313	2.297		-0.7	20.0
2,4,6-Tribromophenol	0.175	0.166		-5.1	20.0
Terphenyl-d14	0.823	0.727		-11.7	20.0
1,4-Dioxane	0.383	0.412		7.6	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q2805
Instrument ID: BNA_N Calibration Date/Time: 08/13/2025 16:35
Lab File ID: BN037598.D Init. Calib. Date(s): 08/12/2025 08/12/2025
EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 16:26 20:05
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.504		-7.4	50.0
Fluoranthene-d10	1.052	0.931		-11.5	50.0
2-Fluorophenol	0.907	0.859		-5.3	50.0
Phenol-d6	1.091	1.022		-6.3	50.0
Nitrobenzene-d5	0.282	0.261		-7.4	50.0
2-Fluorobiphenyl	2.313	2.278		-1.5	50.0
2,4,6-Tribromophenol	0.175	0.157		-10.3	50.0
Terphenyl-d14	0.823	0.804		-2.3	50.0
1,4-Dioxane	0.383	0.389		1.6	50.0

All other compounds must meet a minimum RRF of 0.010.