

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2805

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037597.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	
								Qual		High	RPD
PB169222BSD	1,4-Dioxane	0.4	0.29	ug/L	73	7			70	130	20