

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2806</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>08/13/2025 16:35</u>
Lab File ID:	<u>BN037598.D</u>	Init. Calib. Date(s):	<u>08/12/2025 08/12/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>16:26 20:05</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.504		-7.4	50.0
Fluoranthene-d10	1.052	0.931		-11.5	50.0
2-Fluorophenol	0.907	0.859		-5.3	50.0
Phenol-d6	1.091	1.022		-6.3	50.0
Nitrobenzene-d5	0.282	0.261		-7.4	50.0
2-Fluorobiphenyl	2.313	2.278		-1.5	50.0
2,4,6-Tribromophenol	0.175	0.157		-10.3	50.0
Terphenyl-d14	0.823	0.804		-2.3	50.0
1,4-Dioxane	0.383	0.389		1.6	50.0

All other compounds must meet a minimum RRF of 0.010.