

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081225

Review By	Rahul	Review On	8/13/2025 11:25:23 AM
Supervise By	Jagrut	Supervise On	8/13/2025 11:26:50 AM
SubDirectory	BN081225	HP Acquire Method	BNA_N, 8270_HP Processing Method bn081225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC	SP6846
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037576.D	12 Aug 2025 15:05		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037577.D	12 Aug 2025 15:49	A Fresh Calibration is required	RC/JU	Not Ok
3	SSTDICC0.1	SSTDICC0.1	BN037578.D	12 Aug 2025 16:26		RC/JU	Ok
4	SSTDICC0.2	SSTDICC0.2	BN037579.D	12 Aug 2025 17:03		RC/JU	Ok
5	SSTDICCC0.4	SSTDICCC0.4	BN037580.D	12 Aug 2025 17:39	Compound #20 kept on LR	RC/JU	Ok
6	SSTDICC0.8	SSTDICC0.8	BN037581.D	12 Aug 2025 18:16		RC/JU	Ok
7	SSTDICC1.6	SSTDICC1.6	BN037582.D	12 Aug 2025 18:52		RC/JU	Ok
8	SSTDICC3.2	SSTDICC3.2	BN037583.D	12 Aug 2025 19:29		RC/JU	Ok
9	SSTDICC5.0	SSTDICC5.0	BN037584.D	12 Aug 2025 20:05	Comp #20,23,24 removed from 5ppm	RC/JU	Ok
10	SSTDICV0.4	ICVBN081225	BN037585.D	12 Aug 2025 20:42		RC/JU	Ok
11	PB169094BL	PB169094BL	BN037586.D	12 Aug 2025 21:55		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081325

Review By	Rahul	Review On	8/14/2025 10:14:13 AM
Supervise By	Jagrut	Supervise On	8/15/2025 3:48:33 PM
SubDirectory	BN081325	HP Acquire Method	BNA_N, 8270_HP Processing Method bn081225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6830, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037587.D	13 Aug 2025 09:48		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037588.D	13 Aug 2025 10:28		RC/JU	Ok
3	PB169222BL	PB169222BL	BN037589.D	13 Aug 2025 11:05		RC/JU	Ok
4	Q2805-01	RW8-SP100-20250807	BN037590.D	13 Aug 2025 11:42		RC/JU	Ok
5	Q2805-02	RW8-SP303-20250807	BN037591.D	13 Aug 2025 12:18		RC/JU	Ok
6	Q2806-01	RW7-SP100-20250807	BN037592.D	13 Aug 2025 12:55		RC/JU	Ok
7	Q2806-02	RW7-SP201-20250807	BN037593.D	13 Aug 2025 13:31		RC/JU	Ok
8	Q2806-03	RW7-SP303-20250807	BN037594.D	13 Aug 2025 14:08		RC/JU	Ok
9	Q2825-01	TW1	BN037595.D	13 Aug 2025 14:45		RC/JU	Ok
10	PB169222BS	PB169222BS	BN037596.D	13 Aug 2025 15:22		RC/JU	Ok
11	PB169222BSD	PB169222BSD	BN037597.D	13 Aug 2025 15:58		RC/JU	Ok
12	SSTDCCC0.4	SSTDCCC0.4EC	BN037598.D	13 Aug 2025 16:35		RC/JU	Ok

M : Manual Integration