

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2806

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037596.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB169222BS	1,4-Dioxane	0.4	0.27	ug/L	68		*		70	130