



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Kleinfelder**

**Project Name: Girard School - PA**

**Project # N/A**

**Order ID # Q2807**

**Test Name: Ammonia, Anions Group1, Hexavalent Chromium, Trivalent Chromium**

### **A. Number of Samples and Date of Receipt:**

3 Solid samples were received on 08/08/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group1, Hexavalent Chromium, Trivalent Chromium. This data package contains results for Ammonia, Anions Group1, Hexavalent Chromium, Trivalent Chromium.

### **C. Analytical Techniques:**

The analysis of Trivalent Chromium was based on method 6010D, The analysis of Hexavalent Chromium was based on method 7196A, The analysis of Anions Group1 was based on method 9056A and The analysis of Ammonia was based on method SM4500-NH3.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (COMP-4MS) analysis met criteria for all compounds except for Anions Group1(Fluoride) due to sample matrix interference.

The Matrix Spike Duplicate (COMP-4MSD) analysis met criteria for all compounds except for Anions Group1(Fluoride) due to sample matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_