

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2807

Project: Girard School - PA

RunNo.: LB136749

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.3	10	103	90-110	08/06/2025
Chloride	mg/L	3.2	3	107	90-110	08/06/2025
Fluoride	mg/L	2	2	100	90-110	08/06/2025
Nitrite	mg/L	3.1	3	103	90-110	08/06/2025
Nitrate	mg/L	2.6	2.5	104	90-110	08/06/2025
Sulfate	mg/L	15.4	15	103	90-110	08/06/2025
Orthophosphate as P	mg/L	5	5	100	90-110	08/06/2025
Sample ID: CCV1						
Bromide	mg/L	10.1	10	101	90-110	08/08/2025
Chloride	mg/L	3	3	100	90-110	08/08/2025
Fluoride	mg/L	2	2	100	90-110	08/08/2025
Nitrite	mg/L	3	3	100	90-110	08/08/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/08/2025
Sulfate	mg/L	15	15	100	90-110	08/08/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	08/08/2025
Sample ID: CCV2						
Bromide	mg/L	10.1	10	101	90-110	08/08/2025
Chloride	mg/L	3	3	100	90-110	08/08/2025
Fluoride	mg/L	2	2	100	90-110	08/08/2025
Nitrite	mg/L	3	3	100	90-110	08/08/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/08/2025
Sulfate	mg/L	14.9	15	99	90-110	08/08/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	08/08/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2807

Project: Girard School - PA

RunNo.: LB136775

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	08/11/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	08/11/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	08/11/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	08/11/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2807

Project: Girard School - PA

RunNo.: LB136784

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	08/12/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	08/12/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.97	1	97	90-110	08/12/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	08/12/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.99	1	99	90-110	08/12/2025