

Duplicate Sample Summary

Client: Kleinfelder Project: Girard School - PA Client ID: COMP-1DUP	SDG No.: Q2807 Sample ID: Q2795-01 Percent Solids for Spike Sample: 82.5
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.083	U	0.083	U	1	0		08/11/2025
Ammonia as N	mg/Kg	+/-20	2.60	U	2.60	U	1	0		08/12/2025

Duplicate Sample Summary

Client: Kleinfelder Project: Girard School - PA Client ID: COMP-1MSD	SDG No.: Q2807 Sample ID: Q2795-01 Percent Solids for Spike Sample: 82.5
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	55.3		56.0		1	1		08/12/2025

Duplicate Sample Summary

Client: Kleinfelder	SDG No.: Q2807
Project: Girard School - PA	Sample ID: Q2807-01
Client ID: COMP-4MSD	Percent Solids for Spike Sample: 83.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Bromide	mg/Kg	+/-15	248		247		1	0		08/08/2025
Nitrate	mg/Kg	+/-15	62.6		62.3		1	0		08/08/2025
Nitrite	mg/Kg	+/-15	73.7		73.5		1	0		08/08/2025
Chloride	mg/Kg	+/-15	122		121		1	1		08/08/2025
Fluoride	mg/Kg	+/-15	47.7		47.2		1	1		08/08/2025
Sulfate	mg/Kg	+/-15	375		373		1	1		08/08/2025
Orthophosphate as P	mg/Kg	+/-15	95.5		92.0		1	4		08/08/2025