

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder

SDG No.: Q2807

Contract: POWE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169179BS Mercury	mg/Kg	0.28	0.26		94	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder

SDG No.: Q2807

Contract: POWE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169181BS							
Aluminum	mg/Kg	100	98.7		99	80 - 120	P
Antimony	mg/Kg	40.0	42.0		105	80 - 120	P
Arsenic	mg/Kg	40.0	39.7		99	80 - 120	P
Barium	mg/Kg	10.0	9.19		92	80 - 120	P
Beryllium	mg/Kg	10.0	9.20		92	80 - 120	P
Cadmium	mg/Kg	10.0	9.07		91	80 - 120	P
Calcium	mg/Kg	50.0	56.8	J	114	80 - 120	P
Chromium	mg/Kg	20.0	20.3		102	80 - 120	P
Cobalt	mg/Kg	10.0	9.39		94	80 - 120	P
Copper	mg/Kg	15.0	15.1		101	80 - 120	P
Iron	mg/Kg	150	152		101	80 - 120	P
Lead	mg/Kg	50.0	45.5		91	80 - 120	P
Magnesium	mg/Kg	100	98.2	J	98	80 - 120	P
Manganese	mg/Kg	10.0	9.89		99	80 - 120	P
Nickel	mg/Kg	25.0	23.7		95	80 - 120	P
Potassium	mg/Kg	500	479		96	80 - 120	P
Selenium	mg/Kg	100	99.8		100	80 - 120	P
Silver	mg/Kg	3.8	3.63		96	80 - 120	P
Sodium	mg/Kg	150	142		95	80 - 120	P
Thallium	mg/Kg	100	88.3		88	80 - 120	P
Vanadium	mg/Kg	15.0	14.9		99	80 - 120	P
Zinc	mg/Kg	10.0	10.6		106	80 - 120	P