

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2807

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025 07/23/2025

Continuing Calib Time: 09:29

Initial Calibration Time(s): 11:32 19:47

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.01
Decachlorobiphenyl	8.68	8.69	8.59	8.79	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2807

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025 07/23/2025

Continuing Calib Time: 09:29

Initial Calibration Time(s): 11:32 19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01