

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2807

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 07/23/2025 07/23/2025

Continuing Calib Time: 02:34

Initial Calibration Time(s): 11:32 19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57                   | 3.77 | 0.00       |
| Decachlorobiphenyl   | 8.69       | 8.69      | 8.59                   | 8.79 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** POWE02

**Lab Code:** ACE

**SDG NO.:** Q2807

**Continuing Calib Date:** 08/13/2025

**Initial Calibration Date(s):** 07/23/2025 07/23/2025

**Continuing Calib Time:** 02:34

**Initial Calibration Time(s):** 11:32 19:47

**GC Column:** ZB-MR2 **ID:** 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56                   | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.63       | 8.64      | 8.54                   | 8.74 | 0.01       |