

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112829.D **Time Analyzed:** 09:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.752	4.655	4.855	562.790	500.000	12.6
Aroclor-1016-2	4.771	4.674	4.874	564.160	500.000	12.8
Aroclor-1016-3	4.827	4.731	4.931	549.990	500.000	10.0
Aroclor-1016-4	4.947	4.851	5.051	551.230	500.000	10.2
Aroclor-1016-5	5.203	5.108	5.308	564.730	500.000	12.9
Aroclor-1260-1	6.239	6.145	6.345	544.060	500.000	8.8
Aroclor-1260-2	6.429	6.333	6.533	527.450	500.000	5.5
Aroclor-1260-3	6.794	6.700	6.900	548.500	500.000	9.7
Aroclor-1260-4	7.053	6.959	7.159	505.950	500.000	1.2
Aroclor-1260-5	7.296	7.203	7.403	494.820	500.000	-1.0
Decachlorobiphenyl	8.685	8.593	8.793	48.270	50.000	-3.5
Tetrachloro-m-xylene	3.666	3.569	3.769	59.620	50.000	19.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112829.D **Time Analyzed:** 09:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.732	4.638	4.838	528.370	500.000	5.7
Aroclor-1016-2	4.750	4.656	4.856	524.210	500.000	4.8
Aroclor-1016-3	4.925	4.831	5.031	506.610	500.000	1.3
Aroclor-1016-4	4.968	4.873	5.073	510.430	500.000	2.1
Aroclor-1016-5	5.180	5.085	5.285	526.910	500.000	5.4
Aroclor-1260-1	6.207	6.113	6.313	527.430	500.000	5.5
Aroclor-1260-2	6.395	6.302	6.502	530.180	500.000	6.0
Aroclor-1260-3	6.546	6.453	6.653	543.660	500.000	8.7
Aroclor-1260-4	7.015	6.923	7.123	544.260	500.000	8.9
Aroclor-1260-5	7.257	7.164	7.364	563.790	500.000	12.8
Decachlorobiphenyl	8.630	8.540	8.740	51.850	50.000	3.7
Tetrachloro-m-xylene	3.659	3.562	3.762	56.290	50.000	12.6