

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112844.D **Time Analyzed:** 18:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.752	4.655	4.855	532.140	500.000	6.4
Aroclor-1016-2	4.771	4.674	4.874	528.650	500.000	5.7
Aroclor-1016-3	4.828	4.731	4.931	509.980	500.000	2.0
Aroclor-1016-4	4.947	4.851	5.051	517.910	500.000	3.6
Aroclor-1016-5	5.203	5.108	5.308	543.290	500.000	8.7
Aroclor-1260-1	6.239	6.145	6.345	563.200	500.000	12.6
Aroclor-1260-2	6.429	6.333	6.533	524.250	500.000	4.9
Aroclor-1260-3	6.795	6.700	6.900	451.160	500.000	-9.8
Aroclor-1260-4	7.054	6.959	7.159	424.560	500.000	-15.1
Aroclor-1260-5	7.297	7.203	7.403	400.740	500.000	-19.9
Decachlorobiphenyl	8.687	8.593	8.793	48.740	50.000	-2.5
Tetrachloro-m-xylene	3.667	3.569	3.769	54.360	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112844.D **Time Analyzed:** 18:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.638	4.838	488.920	500.000	-2.2
Aroclor-1016-2	4.751	4.656	4.856	491.740	500.000	-1.7
Aroclor-1016-3	4.926	4.831	5.031	481.850	500.000	-3.6
Aroclor-1016-4	4.969	4.873	5.073	484.840	500.000	-3.0
Aroclor-1016-5	5.180	5.085	5.285	503.530	500.000	0.7
Aroclor-1260-1	6.208	6.113	6.313	481.150	500.000	-3.8
Aroclor-1260-2	6.396	6.302	6.502	475.820	500.000	-4.8
Aroclor-1260-3	6.548	6.453	6.653	462.550	500.000	-7.5
Aroclor-1260-4	7.017	6.923	7.123	468.560	500.000	-6.3
Aroclor-1260-5	7.259	7.164	7.364	480.810	500.000	-3.8
Decachlorobiphenyl	8.633	8.540	8.740	57.700	50.000	15.4
Tetrachloro-m-xylene	3.659	3.562	3.762	51.730	50.000	3.5