

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112850.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.750	4.655	4.855	651.000	500.000	30.2
Aroclor-1016-2	4.769	4.674	4.874	549.480	500.000	9.9
Aroclor-1016-3	4.826	4.731	4.931	565.780	500.000	13.2
Aroclor-1016-4	4.945	4.851	5.051	571.540	500.000	14.3
Aroclor-1016-5	5.202	5.108	5.308	574.860	500.000	15.0
Aroclor-1260-1	6.238	6.145	6.345	579.220	500.000	15.8
Aroclor-1260-2	6.428	6.333	6.533	547.500	500.000	9.5
Aroclor-1260-3	6.794	6.700	6.900	528.750	500.000	5.8
Aroclor-1260-4	7.053	6.959	7.159	498.850	500.000	-0.2
Aroclor-1260-5	7.296	7.203	7.403	553.130	500.000	10.6
Decachlorobiphenyl	8.686	8.593	8.793	46.550	50.000	-6.9
Tetrachloro-m-xylene	3.665	3.569	3.769	60.080	50.000	20.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112850.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.731	4.638	4.838	532.580	500.000	6.5
Aroclor-1016-2	4.749	4.656	4.856	524.330	500.000	4.9
Aroclor-1016-3	4.924	4.831	5.031	531.470	500.000	6.3
Aroclor-1016-4	4.966	4.873	5.073	529.880	500.000	6.0
Aroclor-1016-5	5.179	5.085	5.285	523.970	500.000	4.8
Aroclor-1260-1	6.206	6.113	6.313	535.240	500.000	7.0
Aroclor-1260-2	6.394	6.302	6.502	538.810	500.000	7.8
Aroclor-1260-3	6.545	6.453	6.653	559.410	500.000	11.9
Aroclor-1260-4	7.015	6.923	7.123	530.130	500.000	6.0
Aroclor-1260-5	7.257	7.164	7.364	549.900	500.000	10.0
Decachlorobiphenyl	8.631	8.540	8.740	52.550	50.000	5.1
Tetrachloro-m-xylene	3.658	3.562	3.762	56.390	50.000	12.8