

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: POWE02
SDG NO.: Q2807

Calibration Date(s): 07/31/2025 07/31/2025
Calibration Times: 11:47 12:41

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD089688.D</u>	CF 075 =	<u>PD089689.D</u>		
CF 050 =		<u>PD089690.D</u>	CF 025 =	<u>PD089691.D</u>	CF 005 =	<u>PD089692.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3170910000	3047700000	2999940000	2847680000	2907880000	2994820000	4
4,4'-DDE	4003810000	3817790000	3749630000	3527790000	3549710000	3729750000	5
4,4'-DDT	3548070000	3416270000	3370890000	3205660000	3260320000	3360240000	4
Aldrin	4875360000	4656550000	4601860000	4327300000	4363460000	4564910000	5
Decachlorobiphenyl	3483240000	3499950000	3619460000	3726220000	4252410000	3716260000	8
Dieldrin	4482180000	4304270000	4237810000	4013690000	4079190000	4223430000	4
Tetrachloro-m-xylene	2500060000	2427080000	2456600000	2403480000	2545710000	2466590000	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2807

Instrument ID: ECD_D

Calibration Date(s): 07/31/2025 07/31/2025

Calibration Times: 11:47 12:41

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD089688.D</u>	CF 075 =	<u>PD089689.D</u>		
CF 050 =		<u>PD089690.D</u>	CF 025 =	<u>PD089691.D</u>	CF 005 =	<u>PD089692.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	17134400000	16787300000	17354900000	17662200000	19224300000	17632600000	5
4,4'-DDE	20430900000	20064600000	20732800000	21183800000	23273600000	21137100000	6
4,4'-DDT	18425400000	17982900000	18452700000	18412200000	18590200000	18372700000	1
Aldrin	21657100000	21218800000	21871300000	22261000000	24245900000	22250800000	5
Decachlorobiphenyl	19095200000	18752300000	19614700000	20093800000	22833900000	20078000000	8
Dieldrin	20963500000	20551400000	21238000000	21609600000	23832300000	21639000000	6
Tetrachloro-m-xylene	15162800000	14913400000	15457700000	15824900000	17830500000	15837800000	7