

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2807

Continuing Calib Date: 08/11/2025

Initial Calibration Date(s): 07/31/2025 07/31/2025

Continuing Calib Time: 19:50

Initial Calibration Time(s): 11:47 12:41

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.07       | 9.07      | 8.97                   | 9.17 | 0.00       |
| Tetrachloro-m-xylene | 3.55       | 3.55      | 3.45                   | 3.65 | 0.00       |
| Aldrin               | 5.27       | 5.27      | 5.17                   | 5.37 | 0.00       |
| Dieldrin             | 6.34       | 6.35      | 6.25                   | 6.45 | 0.01       |
| 4,4'-DDE             | 6.19       | 6.19      | 6.09                   | 6.29 | 0.00       |
| 4,4'-DDD             | 6.70       | 6.70      | 6.60                   | 6.80 | 0.00       |
| 4,4'-DDT             | 7.02       | 7.02      | 6.92                   | 7.12 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2807

Continuing Calib Date: 08/11/2025

Initial Calibration Date(s): 07/31/2025 07/31/2025

Continuing Calib Time: 19:50

Initial Calibration Time(s): 11:47 12:41

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 8.07       | 8.07      | 7.97                   | 8.17 | 0.00       |
| Tetrachloro-m-xylene | 2.88       | 2.88      | 2.78                   | 2.98 | 0.00       |
| Aldrin               | 4.37       | 4.37      | 4.27                   | 4.47 | 0.01       |
| Dieldrin             | 5.51       | 5.51      | 5.41                   | 5.61 | 0.00       |
| 4,4'-DDE             | 5.37       | 5.37      | 5.27                   | 5.47 | 0.00       |
| 4,4'-DDD             | 5.92       | 5.93      | 5.83                   | 6.03 | 0.01       |
| 4,4'-DDT             | 6.18       | 6.18      | 6.08                   | 6.28 | 0.00       |