

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089843.D **Time Analyzed:** 19:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.702	6.604	6.804	49.920	50.000	-0.2
4,4'-DDE	6.192	6.094	6.294	48.440	50.000	-3.1
4,4'-DDT	7.017	6.919	7.119	43.190	50.000	-13.6
Aldrin	5.267	5.169	5.369	50.390	50.000	0.8
Decachlorobiphenyl	9.067	8.971	9.171	44.890	50.000	-10.2
Dieldrin	6.343	6.245	6.445	48.220	50.000	-3.6
Tetrachloro-m-xylene	3.548	3.448	3.648	49.870	50.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089843.D **Time Analyzed:** 19:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.924	5.825	6.025	60.370	50.000	20.7
4,4'-DDE	5.370	5.271	5.471	59.310	50.000	18.6
4,4'-DDT	6.178	6.079	6.279	56.010	50.000	12.0
Aldrin	4.365	4.265	4.465	59.550	50.000	19.1
Decachlorobiphenyl	8.066	7.968	8.168	53.090	50.000	6.2
Dieldrin	5.508	5.409	5.609	58.960	50.000	17.9
Tetrachloro-m-xylene	2.878	2.778	2.978	58.790	50.000	17.6