

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089852.D **Time Analyzed:** 22:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.701	6.604	6.804	46.190	50.000	-7.6
4,4'-DDE	6.192	6.094	6.294	45.170	50.000	-9.7
4,4'-DDT	7.016	6.919	7.119	40.170	50.000	-19.7
Aldrin	5.267	5.169	5.369	49.800	50.000	-0.4
Decachlorobiphenyl	9.067	8.971	9.171	42.060	50.000	-15.9
Dieldrin	6.343	6.245	6.445	44.720	50.000	-10.6
Tetrachloro-m-xylene	3.547	3.448	3.648	51.160	50.000	2.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089852.D **Time Analyzed:** 22:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.924	5.825	6.025	55.240	50.000	10.5
4,4'-DDE	5.370	5.271	5.471	56.680	50.000	13.4
4,4'-DDT	6.178	6.079	6.279	48.610	50.000	-2.8
Aldrin	4.365	4.265	4.465	59.560	50.000	19.1
Decachlorobiphenyl	8.065	7.968	8.168	40.150	50.000	-19.7
Dieldrin	5.508	5.409	5.609	55.340	50.000	10.7
Tetrachloro-m-xylene	2.876	2.778	2.978	59.860	50.000	19.7