

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089880.D **Time Analyzed:** 12:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.604	6.804	46.470	50.000	-7.1
4,4'-DDE	6.202	6.094	6.294	46.920	50.000	-6.2
4,4'-DDT	7.026	6.919	7.119	45.880	50.000	-8.2
Aldrin	5.276	5.169	5.369	47.220	50.000	-5.6
Decachlorobiphenyl	9.078	8.971	9.171	44.930	50.000	-10.1
Dieldrin	6.352	6.245	6.445	46.360	50.000	-7.3
Tetrachloro-m-xylene	3.554	3.448	3.648	47.050	50.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089880.D **Time Analyzed:** 12:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	5.927	5.825	6.025	56.590	50.000	13.2
4,4'-DDE	5.373	5.271	5.471	56.850	50.000	13.7
4,4'-DDT	6.181	6.079	6.279	58.380	50.000	16.8
Aldrin	4.366	4.265	4.465	56.210	50.000	12.4
Decachlorobiphenyl	8.070	7.968	8.168	53.030	50.000	6.1
Dieldrin	5.511	5.409	5.609	56.360	50.000	12.7
Tetrachloro-m-xylene	2.878	2.778	2.978	57.060	50.000	14.1