

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089886.D **Time Analyzed:** 17:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.711	6.604	6.804	46.040	50.000	-7.9
4,4'-DDE	6.202	6.094	6.294	46.890	50.000	-6.2
4,4'-DDT	7.027	6.919	7.119	44.450	50.000	-11.1
Aldrin	5.277	5.169	5.369	48.350	50.000	-3.3
Decachlorobiphenyl	9.079	8.971	9.171	43.600	50.000	-12.8
Dieldrin	6.353	6.245	6.445	46.440	50.000	-7.1
Tetrachloro-m-xylene	3.556	3.448	3.648	47.750	50.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2807
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089886.D **Time Analyzed:** 17:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	5.928	5.825	6.025	58.940	50.000	17.9
4,4'-DDE	5.373	5.271	5.471	59.070	50.000	18.1
4,4'-DDT	6.182	6.079	6.279	59.760	50.000	19.5
Aldrin	4.367	4.265	4.465	59.180	50.000	18.4
Decachlorobiphenyl	8.070	7.968	8.168	50.570	50.000	1.1
Dieldrin	5.511	5.409	5.609	58.700	50.000	17.4
Tetrachloro-m-xylene	2.878	2.778	2.978	57.730	50.000	15.5