

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081125\
 Data File : PD089836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2025 17:46
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 01:21:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD073125.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 01 07:03:58 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.877	123.0E6	923.0E6	49.872	58.278
28)	SA Decachlor...	9.072	8.067	165.4E6	1047.8E6	44.511	52.185
Target Compounds							
2)	A alpha-BHC	4.001	3.389	256.8E6	1445.4E6	51.836	59.471
3)	MA gamma-BHC...	4.332	3.726	243.0E6	1326.6E6	51.127	58.897
4)	MA Heptachlor	4.930	4.079	230.0E6	1300.5E6	47.795	56.033
5)	MB Aldrin	5.271	4.364	230.6E6	1314.5E6	50.515	59.077
6)	B beta-BHC	4.518	4.022	94286171	564.9E6	51.113	58.540
7)	B delta-BHC	4.766	4.258	234.6E6	1337.4E6	51.145	58.877
8)	B Heptachlo...	5.691	4.868	203.8E6	1190.1E6	48.748	57.912
9)	A Endosulfan I	6.075	5.242	191.9E6	1107.9E6	48.404	56.675
10)	B gamma-Chl...	5.946	5.121	204.8E6	1278.6E6	49.013	59.048
11)	B alpha-Chl...	6.027	5.186	205.0E6	1222.9E6	48.760	58.731
12)	B 4,4'-DDE	6.196	5.370	180.5E6	1241.7E6	48.392	58.747
13)	MA Dieldrin	6.347	5.508	203.8E6	1259.6E6	48.249	58.208
14)	MA Endrin	6.575	5.785	160.3E6	1112.7E6	44.797	55.842
15)	B Endosulfa...	6.787	6.076	171.0E6	1083.4E6	46.601	57.628
16)	A 4,4'-DDD	6.706	5.925	148.8E6	1057.2E6	49.679	59.958
17)	MA 4,4'-DDT	7.022	6.179	143.0E6	1008.8E6	42.553	54.909 #
18)	B Endrin al...	6.915	6.254	117.7E6	748.1E6	47.173	57.898
19)	B Endosulfa...	7.150	6.478	159.2E6	1036.9E6	45.874	55.852
20)	A Methoxychlor	7.494	6.749	76343951	533.7E6	42.699	54.489 #
21)	B Endrin ke...	7.631	6.987	171.9E6	1173.8E6	46.291	58.491 #
22)	Mirex	8.113	7.180	125.2E6	879.0E6	44.905	57.749 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081125\
Data File : PD089836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2025 17:46
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 01:21:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD073125.M
Quant Title : GC Extractables
QLast Update : Fri Aug 01 07:03:58 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

