

Surrogate Summary

SDG No.: **Q2807**

Client: **Kleinfelder**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089685.D	PIBLK-PD089685.D	Decachlorobiphen	1	20	20.9	104		57	171
		Tetrachloro-m-xyl	1	20	19.6	98		61	148
		Decachlorobiphen	2	20	20.9	104		57	171
		Tetrachloro-m-xyl	2	20	20.4	102		61	148
I.BLK-PD089835.D	PIBLK-PD089835.D	Decachlorobiphen	1	20	18.9	94		57	171
		Tetrachloro-m-xyl	1	20	19.4	97		61	148
		Decachlorobiphen	2	20	22.1	110		57	171
		Tetrachloro-m-xyl	2	20	24.2	121		61	148
PB169187BL	PB169187BL	Decachlorobiphen	1	20	18.7	94		20	144
		Tetrachloro-m-xyl	1	20	18.6	93		19	148
		Decachlorobiphen	2	20	21.8	109		20	144
		Tetrachloro-m-xyl	2	20	23.6	118		19	148
PB169187BS	PB169187BS	Decachlorobiphen	1	20	21.1	105		20	144
		Tetrachloro-m-xyl	1	20	21.0	105		19	148
		Decachlorobiphen	2	20	24.7	124		20	144
		Tetrachloro-m-xyl	2	20	26.5	133		19	148
I.BLK-PD089841.D	PIBLK-PD089841.D	Decachlorobiphen	1	20	19.1	95		57	171
		Tetrachloro-m-xyl	1	20	19.7	98		61	148
		Decachlorobiphen	2	20	22.4	112		57	171
		Tetrachloro-m-xyl	2	20	24.3	122		61	148
Q2807-01	COMP-4	Decachlorobiphen	1	20	20.9	104		20	144
		Tetrachloro-m-xyl	1	20	24.1	121		19	148
		Decachlorobiphen	2	20	24.5	122		20	144
		Tetrachloro-m-xyl	2	20	29.3	146		19	148
Q2807-01MS	COMP-4MS	Decachlorobiphen	1	20	15.8	79		20	144
		Tetrachloro-m-xyl	1	20	17.4	87		19	148
		Decachlorobiphen	2	20	18.2	91		20	144
		Tetrachloro-m-xyl	2	20	21.9	109		19	148
Q2807-01MSD	COMP-4MSD	Decachlorobiphen	1	20	15.4	77		20	144
		Tetrachloro-m-xyl	1	20	17.5	88		19	148
		Decachlorobiphen	2	20	18.2	91		20	144
		Tetrachloro-m-xyl	2	20	22.1	111		19	148
Q2807-03	COMP-6	Decachlorobiphen	1	20	13.2	66		20	144
		Tetrachloro-m-xyl	1	20	15.3	76		19	148
		Decachlorobiphen	2	20	15.1	76		20	144
		Tetrachloro-m-xyl	2	20	18.8	94		19	148
I.BLK-PD089851.D	PIBLK-PD089851.D	Decachlorobiphen	1	20	16.5	83		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	13.0	65		57	171
		Tetrachloro-m-xyl	2	20	24.0	120		61	148
I.BLK-PD089878.D	PIBLK-PD089878.D	Decachlorobiphen	1	20	19.5	97		57	171

Surrogate Summary

SDG No.: Q2807

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089878.D	PIBLK-PD089878.D	Tetrachloro-m-xyl	1	20	19.3	97		61	148
		Decachlorobiphen	2	20	22.9	114		57	171
Q2807-02	COMP-5	Tetrachloro-m-xyl	2	20	24.8	124		61	148
		Decachlorobiphen	1	20	24.3	122		20	144
		Tetrachloro-m-xyl	1	20	26.1	130		19	148
		Decachlorobiphen	2	20	28.8	144		20	144
		Tetrachloro-m-xyl	2	20	32.5	163	*	19	148
I.BLK-PD089885.D	PIBLK-PD089885.D	Decachlorobiphen	1	20	19.3	96		57	171
		Tetrachloro-m-xyl	1	20	19.3	97		61	148
		Decachlorobiphen	2	20	21.5	108		57	171
		Tetrachloro-m-xyl	2	20	24.4	122		61	148