

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF081225

SequenceID :	<u>BF081225</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)		_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)		_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
5. GC/MS Calibration Met:		_____	_____	_____✓_____
a. Initial calibration Meet Criteria		_____	_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____✓_____
The %RSD is greater than 20% for certain compounds in the Initial Calibration (Method 8270-BF081225.M) for 2-Nitrophenol, Hexachlorocyclopentadiene, 2-Nitroaniline, 2,6-Dinitrotoluene, 4-Nitrophenol, 2,4-Dinitrotoluene, 4,6-Dinitro-2-methylephenol,				
b. Continuous Calibration(CCC) Meet Criteria		_____	_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____✓_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:		_____	_____✓_____	_____
a. B/N Fraction				
d. Acid Fraction				

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

____✓

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

____✓

a. B/N Fraction

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

____✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

____✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

____✓

ADDITIONAL COMMENTS:

Ashish

Analyst

08/12/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF081225

SequenceID : BF081225 NA NO YES