

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081225

Review By	Jagrut	Review On	8/13/2025 11:12:27 AM
Supervise By	Rahul	Supervise On	8/13/2025 3:44:43 PM
SubDirectory	BF081225	HP Acquire Method	BNA_F
		HP Processing Method	bf081225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13170,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143342.D	12 Aug 2025 08:34		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF143343.D	12 Aug 2025 09:03		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF143344.D	12 Aug 2025 09:34		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF143345.D	12 Aug 2025 10:04		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF143346.D	12 Aug 2025 10:35		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF143347.D	12 Aug 2025 11:06	Compound #26,41,48,51,56,57,65,70,80,84, 85 Kept on LR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF143348.D	12 Aug 2025 11:37	Compound #32,54 kept on QR	RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF143349.D	12 Aug 2025 12:07		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF143350.D	12 Aug 2025 12:38	Compound#09 removed from 80 ppm	RC/JU	Ok
10	SSTDICV040	ICVBF081225	BF143351.D	12 Aug 2025 13:52		RC/JU	Ok
11	PB169189BL	PB169189BL	BF143352.D	12 Aug 2025 14:22		RC/JU	Ok
12	PB169189BS	PB169189BS	BF143353.D	12 Aug 2025 14:53	Compound#44,51,59,62,74 failing high	RC/JU	Ok,M
13	Q2807-02	COMP-5	BF143354.D	12 Aug 2025 15:33		RC/JU	Ok,M
14	Q2807-01	COMP-4	BF143355.D	12 Aug 2025 16:02		RC/JU	Ok
15	Q2807-03	COMP-6	BF143356.D	12 Aug 2025 16:32		RC/JU	Ok,M
16	Q2807-03MS	COMP-6MS	BF143357.D	12 Aug 2025 17:02		RC/JU	Ok
17	Q2807-03MSD	COMP-6MSD	BF143358.D	12 Aug 2025 17:32		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081225

Review By	Jagrut	Review On	8/13/2025 11:12:27 AM		
Supervise By	Rahul	Supervise On	8/13/2025 3:44:43 PM		
SubDirectory	BF081225	HP Acquire Method	BNA_F	HP Processing Method	bf081225
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13170,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2830-07	BIN0009-DRIVEWAY-T	BF143359.D	12 Aug 2025 18:02		RC/JU	Ok
19	Q2830-01	BIN0009-DRIVEWAY-T	BF143360.D	12 Aug 2025 18:32		RC/JU	Ok
20	Q2830-03	BIN0009-DRIVEWAY-T	BF143361.D	12 Aug 2025 19:02		RC/JU	Ok
21	Q2830-05	BIN0009-DRIVEWAY-T	BF143362.D	12 Aug 2025 19:32		RC/JU	Ok
22	Q2831-01	VNJ-238	BF143363.D	12 Aug 2025 20:02		RC/JU	Ok,M

M : Manual Integration