

Surrogate Summary

SW-846

SDG No.: Q2807

Client: Kleinfelder

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169189BL	PB169189BL	Nitrobenzene-d5	100	95.5	96		18	107
		2-Fluorobiphenyl	100	84.7	85		20	109
		Terphenyl-d14	100	97.6	98		10	105
PB169189BS	PB169189BS	Nitrobenzene-d5	100	94.3	94		18	107
		2-Fluorobiphenyl	100	85.3	85		20	109
		Terphenyl-d14	100	87.8	88		10	105
Q2807-01	COMP-4	Nitrobenzene-d5	100	68.0	68		18	107
		2-Fluorobiphenyl	100	68.4	68		20	109
		Terphenyl-d14	100	54.3	54		10	105
Q2807-02	COMP-5	Nitrobenzene-d5	100	49.3	49		18	107
		2-Fluorobiphenyl	100	48.9	49		20	109
		Terphenyl-d14	100	38.8	39		10	105
Q2807-03	COMP-6	Nitrobenzene-d5	100	74.4	74		18	107
		2-Fluorobiphenyl	100	71.3	71		20	109
		Terphenyl-d14	100	55.8	56		10	105
Q2807-03MS	COMP-6MS	Nitrobenzene-d5	100	70.5	70		18	107
		2-Fluorobiphenyl	100	65.9	66		20	109
		Terphenyl-d14	100	53.3	53		10	105
Q2807-03MSD	COMP-6MSD	Nitrobenzene-d5	100	71.1	71		18	107
		2-Fluorobiphenyl	100	64.5	65		20	109
		Terphenyl-d14	100	59.1	59		10	105