

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2807

Client ID : SSTDICCC040

Date Analyzed: 08/12/2025

Lab File ID: BF143347.D

Time Analyzed: 11:06

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	167361	6.934	655032	8.21	345532	9.97
UPPER LIMIT	334722	7.434	1310060	8.71	691064	10.469
LOWER LIMIT	83680.5	6.434	327516	7.71	172766	9.469
EPA SAMPLE NO.						
01 PB169189BL	164595	6.93	636089	8.21	350207	9.96
02 PB169189BS	165936	6.93	641337	8.21	345183	9.97
03 COMP-4	146370	6.93	551165	8.21	273111	9.96
04 COMP-5	133341	6.93	496028	8.21	252686	9.96
05 COMP-6	152247	6.93	565615	8.21	280798	9.96
06 COMP-6MS	156492	6.93	586819	8.21	297224	9.96
07 COMP-6MSD	182920	6.93	692344	8.21	366844	9.97

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.