

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143342.D
Instrument ID: BNA_F

Contract: POWE02
SDG NO.: Q2807
DFTPP Injection Date: 08/12/2025
DFTPP Injection Time: 08:34

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	29.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.5
365	Greater than 1% of mass 198	2.9
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF143343.D	08/12/2025	09:03
SSTDICC005	SSTDICC005	BF143344.D	08/12/2025	09:34
SSTDICC010	SSTDICC010	BF143345.D	08/12/2025	10:04
SSTDICC020	SSTDICC020	BF143346.D	08/12/2025	10:35
SSTDICCC040	SSTDICCC040	BF143347.D	08/12/2025	11:06
SSTDICC050	SSTDICC050	BF143348.D	08/12/2025	11:37
SSTDICC060	SSTDICC060	BF143349.D	08/12/2025	12:07
SSTDICC080	SSTDICC080	BF143350.D	08/12/2025	12:38
PB169189BL	PB169189BL	BF143352.D	08/12/2025	14:22
PB169189BS	PB169189BS	BF143353.D	08/12/2025	14:53
COMP-5	Q2807-02	BF143354.D	08/12/2025	15:33
COMP-4	Q2807-01	BF143355.D	08/12/2025	16:02
COMP-6	Q2807-03	BF143356.D	08/12/2025	16:32
COMP-6MS	Q2807-03MS	BF143357.D	08/12/2025	17:02
COMP-6MSD	Q2807-03MSD	BF143358.D	08/12/2025	17:32